

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF092421\
 Data File : BF125617.D
 Acq On : 24 Sep 2021 12:05
 Operator : JU/CG
 Sample : SSTDICC020
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC020

Quant Time: Sep 24 13:07:04 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF092421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 24 13:05:07 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.887	152	195957	20.000	ng	0.00
21) Naphthalene-d8	8.169	136	799506	20.000	ng	# 0.00
39) Acenaphthene-d10	9.928	164	425915	20.000	ng	0.00
64) Phenanthrene-d10	11.410	188	752334	20.000	ng	0.00
76) Chrysene-d12	14.045	240	504334	20.000	ng	0.00
86) Perylene-d12	15.521	264	409409	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.498	112	522587	42.172	ng	0.00
7) Phenol-d6	6.504	99	689091	43.130	ng	0.00
23) Nitrobenzene-d5	7.445	82	444733	35.479	ng	0.00
42) 2,4,6-Tribromophenol	10.710	330	174697	41.877	ng	0.00
45) 2-Fluorobiphenyl	9.245	172	1183856	42.023	ng	0.00
79) Terphenyl-d14	12.992	244	1281845	41.043	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	2.669	88	108131	21.643	ng	# 100
3) Pyridine	3.428	79	280445	20.948	ng	# 71
4) n-Nitrosodimethylamine	3.369	42	104143	19.723	ng	# 1
6) Aniline	6.545	93	381590	20.745	ng	95
8) 2-Chlorophenol	6.669	128	295038	20.850	ng	98
9) Benzaldehyde	6.440	77	194039	21.721	ng	94
10) Phenol	6.516	94	344977	21.213	ng	90
11) bis(2-Chloroethyl)ether	6.622	93	263681	20.281	ng	95
12) 1,3-Dichlorobenzene	6.834	146	321149	21.116	ng	# 95
13) 1,4-Dichlorobenzene	6.910	146	327633	21.312	ng	96
14) 1,2-Dichlorobenzene	7.063	146	309032	21.524	ng	98
15) Benzyl Alcohol	7.022	79	232024	20.293	ng	93
16) 2,2'-oxybis(1-Chloropr...	7.163	45	335756	18.769	ng	90
17) 2-Methylphenol	7.128	107	231556	20.501	ng	99
18) Hexachloroethane	7.404	117	103781	18.960	ng	93
19) n-Nitroso-di-n-propyla...	7.292	70	189058	19.771	ng	# 69
20) 3+4-Methylphenols	7.281	107	304226	21.662	ng	# 81
22) Acetophenone	7.292	105	388851	21.021	ng	# 90
24) Nitrobenzene	7.463	77	244291	18.391	ng	90
25) Isophorone	7.704	82	524219	19.433	ng	92
26) 2-Nitrophenol	7.781	139	89090	14.895	ng	88
27) 2,4-Dimethylphenol	7.816	122	237793	19.837	ng	97
28) bis(2-Chloroethoxy)met...	7.916	93	302396	19.511	ng	97
29) 2,4-Dichlorophenol	8.022	162	238117	19.650	ng	99
30) 1,2,4-Trichlorobenzene	8.110	180	260355	20.386	ng	97
31) Naphthalene	8.192	128	891130	21.703	ng	98
32) Benzoic acid	7.898	122	112165	15.411	ng	93
33) 4-Chloroaniline	8.234	127	359124	21.338	ng	99
34) Hexachlorobutadiene	8.316	225	143317	19.901	ng	97
35) Caprolactam	8.586	113	74501	22.417	ng	# 81
36) 4-Chloro-3-methylphenol	8.710	107	247840	20.820	ng	97
37) 2-Methylnaphthalene	8.881	142	596854	21.461	ng	99
38) 1-Methylnaphthalene	8.981	142	557205	21.353	ng	99
40) 1,2,4,5-Tetrachloroben...	9.051	216	238047	20.757	ng	98
41) Hexachlorocyclopentadiene	9.039	237	98920	15.424	ng	97
43) 2,4,6-Trichlorophenol	9.157	196	165473	19.151	ng	96

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF092421\
 Data File : BF125617.D
 Acq On : 24 Sep 2021 12:05
 Operator : JU/CG
 Sample : SSTDICC020
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC020

Quant Time: Sep 24 13:07:04 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF092421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 24 13:05:07 2021
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	9.192	196	178600	19.696	ng	90
46) 1,1'-Biphenyl	9.345	154	675209	21.040	ng	96
47) 2-Chloronaphthalene	9.369	162	559354	20.986	ng	97
48) 2-Nitroaniline	9.463	65	101553	17.855	ng	87
49) Acenaphthylene	9.786	152	849749	21.693	ng	99
50) Dimethylphthalate	9.639	163	608799	20.769	ng	99
51) 2,6-Dinitrotoluene	9.704	165	97343	16.692	ng	90
52) Acenaphthene	9.957	154	522263	21.121	ng	97
53) 3-Nitroaniline	9.869	138	117356	18.947	ng #	86
54) 2,4-Dinitrophenol	9.969	184	29171	13.884	ng #	1
55) Dibenzofuran	10.128	168	800771	21.823	ng	94
56) 4-Nitrophenol	10.016	139	98181	19.541	ng	91
57) 2,4-Dinitrotoluene	10.104	165	112861	16.149	ng #	84
58) Fluorene	10.475	166	602876	21.362	ng	99
59) 2,3,4,6-Tetrachlorophenol	10.245	232	137854	20.241	ng #	90
60) Diethylphthalate	10.345	149	588131	20.358	ng	98
61) 4-Chlorophenyl-phenyle...	10.463	204	264836	21.100	ng	90
62) 4-Nitroaniline	10.480	138	125224	22.373	ng #	75
63) Azobenzene	10.622	77	569615	20.292	ng	84
65) 4,6-Dinitro-2-methylph...	10.510	198	47161	12.914	ng #	71
66) n-Nitrosodiphenylamine	10.580	169	536648	19.648	ng	97
67) 4-Bromophenyl-phenylether	10.957	248	161160	18.745	ng	94
68) Hexachlorobenzene	11.022	284	193344	19.954	ng #	84
69) Atrazine	11.104	200	147175	20.037	ng	94
70) Pentachlorophenol	11.210	266	100203	18.139	ng	94
71) Phenanthrene	11.433	178	912569	21.594	ng	98
72) Anthracene	11.486	178	910189	21.607	ng	99
73) Carbazole	11.633	167	906656	22.757	ng	98
74) Di-n-butylphthalate	11.969	149	918595	19.427	ng	99
75) Fluoranthene	12.621	202	956258	23.065	ng	99
77) Benzidine	12.739	184	287906	20.133	ng	98
78) Pyrene	12.851	202	975008	20.630	ng	97
80) Butylbenzylphthalate	13.468	149	320551	16.785	ng	98
81) Benzo(a)anthracene	14.033	228	718629	20.932	ng	99
82) 3,3'-Dichlorobenzidine	13.998	252	209807	19.946	ng	99
83) Chrysene	14.074	228	719224	20.739	ng	97
84) Bis(2-ethylhexyl)phtha...	14.027	149	354351	15.178	ng	98
85) Di-n-octyl phthalate	14.639	149	456639	11.763	ng #	90
87) Indeno(1,2,3-cd)pyrene	16.998	276	569558	19.379	ng	98
88) Benzo(b)fluoranthene	15.086	252	583818	19.498	ng	98
89) Benzo(k)fluoranthene	15.086	252	583818	21.320	ng	98
90) Benzo(a)pyrene	15.457	252	525648	20.171	ng	96
91) Dibenzo(a,h)anthracene	17.015	278	487319	19.564	ng	98
92) Benzo(g,h,i)perylene	17.445	276	482729	19.636	ng	93


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF092421\
 Data File : BF125617.D
 Acq On : 24 Sep 2021 12:05
 Operator : JU/CG
 Sample : SSTDICC020
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC020

Quant Time: Sep 24 13:07:04 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF092421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 24 13:05:07 2021
 Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 20.000 ng
 RT: 6.887 min Scan# 816
 Delta R.T. -0.005 min
 Lab File: BF125617.D
 Acq: 24 Sep 2021 12:05




Instrument :
 BNA_F
 ClientSampleId :
 SSTDIC020


Tgt Ion:152 Resp: 195957
 Ion Ratio Lower Upper
 152 100
 150 160.1 117.6 176.4
 115 63.2 47.1 70.7



#2
 1,4-Dioxane
 Concen: 21.643 ng
 RT: 2.669 min Scan# 99
 Delta R.T. -0.000 min
 Lab File: BF125617.D
 Acq: 24 Sep 2021 12:05

Tgt Ion: 88 Resp: 108131
 Ion Ratio Lower Upper
 88 100
 58 68.2 0.0 0.0#
 43 24.0 0.0 0.0#









#3
Pyridine
Concen: 20.948 ng
RT: 3.428 min Scan# 228
Delta R.T. -0.006 min
Lab File: BF125617.D
Acq: 24 Sep 2021 12:05




Instrument :
BNA_F
ClientSampleId :
SSTDIC020


Tgt Ion: 79 Resp: 280445
Ion Ratio Lower Upper
79 100
52 63.8 70.3 105.5#
51 31.5 44.1 66.1#



#4
n-Nitrosodimethylamine
Concen: 19.723 ng
RT: 3.369 min Scan# 218
Delta R.T. -0.012 min
Lab File: BF125617.D
Acq: 24 Sep 2021 12:05

Tgt Ion: 42 Resp: 104143
Ion Ratio Lower Upper
42 100
74 161.2 59.0 88.6#
44 9.2 0.0 0.0#









#5
2-Fluorophenol
Concen: 42.172 ng
RT: 5.498 min Scan# 580
Delta R.T. -0.006 min
Lab File: BF125617.D
Acq: 24 Sep 2021 12:05




Instrument :
BNA_F
ClientSampleId :
SSTDIC020


Tgt Ion:112 Resp: 522587
Ion Ratio Lower Upper
112 100
64 52.6 44.2 66.2
63 25.9 26.2 39.2#



#6
Aniline
Concen: 20.745 ng
RT: 6.545 min Scan# 758
Delta R.T. -0.006 min
Lab File: BF125617.D
Acq: 24 Sep 2021 12:05

Tgt Ion: 93 Resp: 381590
Ion Ratio Lower Upper
93 100
66 37.1 32.7 49.1
65 19.1 14.2 21.4









#7
Phenol-d6
Concen: 43.130 ng
RT: 6.504 min Scan# 751
Delta R.T. -0.006 min
Lab File: BF125617.D
Acq: 24 Sep 2021 12:05




Instrument :
BNA_F
ClientSampleId :
SSTDIC020


Tgt Ion: 99 Resp: 689091
Ion Ratio Lower Upper
99 100
42 14.5 23.5 35.3#
71 30.4 25.6 38.4



#8
2-Chlorophenol
Concen: 20.850 ng
RT: 6.669 min Scan# 779
Delta R.T. -0.006 min
Lab File: BF125617.D
Acq: 24 Sep 2021 12:05

Tgt Ion: 128 Resp: 295038
Ion Ratio Lower Upper
128 100
130 30.7 10.5 50.5
64 41.9 20.3 60.3









#9
Benzaldehyde
Concen: 21.721 ng
RT: 6.440 min Scan# 740
Delta R.T. -0.000 min
Lab File: BF125617.D
Acq: 24 Sep 2021 12:05




Instrument :
BNA_F
ClientSampleId :
SSTDIC020


Tgt Ion: 77 Resp: 194039
Ion Ratio Lower Upper
77 100
105 108.2 75.9 115.9
106 103.5 84.0 124.0



#10
Phenol
Concen: 21.213 ng
RT: 6.516 min Scan# 753
Delta R.T. -0.012 min
Lab File: BF125617.D
Acq: 24 Sep 2021 12:05

Tgt Ion: 94 Resp: 344977
Ion Ratio Lower Upper
94 100
65 25.5 9.3 49.3
66 32.6 20.3 60.3









#11
bis(2-Chloroethyl)ether
Concen: 20.281 ng
RT: 6.622 min Scan# 771
Delta R.T. -0.000 min
Lab File: BF125617.D
Acq: 24 Sep 2021 12:05




Instrument :
BNA_F
ClientSampleId :
SSTDIC020


Tgt Ion: 93 Resp: 263681
Ion Ratio Lower Upper
93 100
63 68.5 54.4 94.4
95 32.0 11.4 51.4



#12
1,3-Dichlorobenzene
Concen: 21.116 ng
RT: 6.834 min Scan# 807
Delta R.T. -0.000 min
Lab File: BF125617.D
Acq: 24 Sep 2021 12:05

Tgt Ion: 146 Resp: 321149
Ion Ratio Lower Upper
146 100
148 62.7 49.7 74.5
75 24.1 25.2 37.8#





#13

1,4-Dichlorobenzene

Concen: 21.312 ng

RT: 6.910 min Scan# 820

Delta R.T. -0.000 min

Lab File: BF125617.D

Acq: 24 Sep 2021 12:05

Instrument :

BNA_F

ClientSampleId :

SSTDIC020





Tgt Ion:146 Resp: 327633

Ion Ratio Lower Upper

146 100

148 61.6 51.9 77.9

111 37.5 31.8 47.6



#14

1,2-Dichlorobenzene

Concen: 21.524 ng

RT: 7.063 min Scan# 846

Delta R.T. -0.000 min

Lab File: BF125617.D

Acq: 24 Sep 2021 12:05




Tgt Ion:146 Resp: 309032


Ion Ratio Lower Upper

146 100

148 62.6 50.6 75.8

111 40.0 33.5 50.3





#15

Benzyl Alcohol

Concen: 20.293 ng

RT: 7.022 min Scan# 839

Delta R.T. -0.006 min

Lab File: BF125617.D

Acq: 24 Sep 2021 12:05

Tgt Ion: 79 Resp: 232024





Ion	Ratio	Lower	Upper
79	100		
108	83.9	60.4	90.6
77	61.7	51.5	77.3

Instrument :

BNA_F

ClientSampleId :

SSTDIC020



#16

2,2'-oxybis(1-Chloropropane)

Concen: 18.769 ng

RT: 7.163 min Scan# 863




Delta R.T. -0.000 min


Lab File: BF125617.D

Acq: 24 Sep 2021 12:05

Tgt Ion: 45 Resp: 335756

Ion	Ratio	Lower	Upper
45	100		
77	14.1	0.0	30.6
79	10.9	0.0	27.5









#17
2-Methylphenol
Concen: 20.501 ng
RT: 7.128 min Scan# 857
Delta R.T. -0.006 min
Lab File: BF125617.D
Acq: 24 Sep 2021 12:05




Instrument :
BNA_F
ClientSampleId :
SSTDIC020


Tgt Ion:	107	Resp:	231556
Ion	Ratio	Lower	Upper
107	100		
108	112.9	90.6	135.8
77	41.2	35.7	53.5
79	42.1	34.0	51.0



#18
Hexachloroethane
Concen: 18.960 ng
RT: 7.404 min Scan# 904
Delta R.T. -0.000 min
Lab File: BF125617.D
Acq: 24 Sep 2021 12:05


Tgt Ion:	117	Resp:	103781
Ion	Ratio	Lower	Upper
117	100		
119	95.3	73.4	110.2
201	89.9	64.9	97.3








#19
n-Nitroso-di-n-propylamine
Concen: 19.771 ng
RT: 7.292 min Scan# 885
Delta R.T. -0.012 min
Lab File: BF125617.D
Acq: 24 Sep 2021 12:05

Instrument :
BNA_F
ClientSampleId :
SSTDIC020







Tgt Ion: 70 Resp: 189058
Ion Ratio Lower Upper
70 100
42 51.5 72.2 108.2#
101 9.8 7.8 11.8
130 22.3 21.8 32.6



#20
3+4-Methylphenols
Concen: 21.662 ng
RT: 7.281 min Scan# 883
Delta R.T. -0.006 min
Lab File: BF125617.D
Acq: 24 Sep 2021 12:05

Tgt Ion: 107 Resp: 304226
Ion Ratio Lower Upper
107 100
108 87.7 76.6 116.6
77 42.2 53.3 93.3#
79 25.4 8.7 48.7









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.169 min Scan# 1034
Delta R.T. -0.006 min
Lab File: BF125617.D
Acq: 24 Sep 2021 12:05




Instrument :
BNA_F
ClientSampleId :
SSTDIC020


Tgt Ion:	136	Resp:	799506
Ion Ratio	Lower	Upper	
136	100		
137	10.9	9.8	14.6
54	6.7	6.4	9.6
68	5.7	3.3	4.9



#22
Acetophenone
Concen: 21.021 ng
RT: 7.292 min Scan# 885
Delta R.T. -0.006 min
Lab File: BF125617.D
Acq: 24 Sep 2021 12:05

Tgt Ion:	105	Resp:	388851
Ion Ratio	Lower	Upper	
105	100		
71	7.4	5.5	8.3
51	22.7	26.0	39.0
120	22.1	18.5	27.7









#23
Nitrobenzene-d5
Concen: 35.479 ng
RT: 7.445 min Scan# 911
Delta R.T. -0.006 min
Lab File: BF125617.D
Acq: 24 Sep 2021 12:05




Instrument :
BNA_F
ClientSampleId :
SSTDIC020


Tgt Ion: 82 Resp: 444733
Ion Ratio Lower Upper
82 100
128 42.8 40.6 60.8
54 45.9 49.6 74.4#



#24
Nitrobenzene
Concen: 18.391 ng
RT: 7.463 min Scan# 914
Delta R.T. -0.006 min
Lab File: BF125617.D
Acq: 24 Sep 2021 12:05

Tgt Ion: 77 Resp: 244291
Ion Ratio Lower Upper
77 100
123 45.5 42.3 63.5
65 12.6 12.4 18.6









#25
Isophorone
Concen: 19.433 ng
RT: 7.704 min Scan# 955
Delta R.T. -0.006 min
Lab File: BF125617.D
Acq: 24 Sep 2021 12:05




Instrument :
BNA_F
ClientSampleId :
SSTDIC020


Tgt Ion: 82 Resp: 524219
Ion Ratio Lower Upper
82 100
95 6.7 6.5 9.7
138 18.5 18.4 27.6



#26
2-Nitrophenol
Concen: 14.895 ng
RT: 7.781 min Scan# 968
Delta R.T. -0.006 min
Lab File: BF125617.D
Acq: 24 Sep 2021 12:05

Tgt Ion:139 Resp: 89090
Ion Ratio Lower Upper
139 100
109 26.2 26.0 39.0
65 46.0 31.0 46.4









#27
2,4-Dimethylphenol
Concen: 19.837 ng
RT: 7.816 min Scan# 974
Delta R.T. -0.000 min
Lab File: BF125617.D
Acq: 24 Sep 2021 12:05




Instrument :
BNA_F
ClientSampleId :
SSTDIC020


Tgt Ion:122 Resp: 237793
Ion Ratio Lower Upper
122 100
107 108.4 85.0 127.6
121 57.6 43.3 64.9



#28
bis(2-Chloroethoxy)methane
Concen: 19.511 ng
RT: 7.916 min Scan# 991
Delta R.T. -0.000 min
Lab File: BF125617.D
Acq: 24 Sep 2021 12:05

Tgt Ion: 93 Resp: 302396
Ion Ratio Lower Upper
93 100
95 32.0 26.1 39.1
123 11.9 11.7 17.5









#29
2,4-Dichlorophenol
Concen: 19.650 ng
RT: 8.022 min Scan# 1009
Delta R.T. -0.000 min
Lab File: BF125617.D
Acq: 24 Sep 2021 12:05




Instrument :
BNA_F
ClientSampleId :
SSTDIC020


Tgt Ion:	162	Resp:	238117
Ion Ratio	Lower	Upper	
162	100		
164	63.1	44.5	84.5
98	33.4	13.0	53.0



#30
1,2,4-Trichlorobenzene
Concen: 20.386 ng
RT: 8.110 min Scan# 1024
Delta R.T. -0.006 min
Lab File: BF125617.D
Acq: 24 Sep 2021 12:05

Tgt Ion:	180	Resp:	260355
Ion Ratio	Lower	Upper	
180	100		
182	93.5	76.8	115.2
145	30.3	26.7	40.1









#31
Naphthalene
Concen: 21.703 ng
RT: 8.192 min Scan# 1038
Delta R.T. -0.006 min
Lab File: BF125617.D
Acq: 24 Sep 2021 12:05




Instrument :
BNA_F
ClientSampleId :
SSTDIC020


Tgt Ion:128 Resp: 891130
Ion Ratio Lower Upper
128 100
129 10.5 9.4 14.0
127 12.9 10.9 16.3



#32
Benzoic acid
Concen: 15.411 ng
RT: 7.898 min Scan# 988
Delta R.T. -0.029 min
Lab File: BF125617.D
Acq: 24 Sep 2021 12:05

Tgt Ion:122 Resp: 112165
Ion Ratio Lower Upper
122 100
105 121.6 97.6 137.6
77 90.2 59.6 99.6









#33
4-Chloroaniline
Concen: 21.338 ng
RT: 8.234 min Scan# 1045
Delta R.T. -0.006 min
Lab File: BF125617.D
Acq: 24 Sep 2021 12:05




Instrument :
BNA_F
ClientSampleId :
SSTDIC020


Tgt Ion:127	Resp:	359124
Ion	Ratio	Lower Upper
127	100	
129	31.6	25.7 38.5
65	26.0	21.2 31.8
92	17.9	14.0 21.0



#34
Hexachlorobutadiene
Concen: 19.901 ng
RT: 8.316 min Scan# 1059
Delta R.T. -0.000 min
Lab File: BF125617.D
Acq: 24 Sep 2021 12:05

Tgt Ion:225	Resp:	143317
Ion	Ratio	Lower Upper
225	100	
223	61.9	50.6 76.0
227	66.1	50.6 75.8









#35
Caprolactam
Concen: 22.417 ng
RT: 8.586 min Scan# 1105
Delta R.T. -0.024 min
Lab File: BF125617.D
Acq: 24 Sep 2021 12:05




Instrument :
BNA_F
ClientSampleId :
SSTDIC020


Tgt Ion:113 Resp: 74501
Ion Ratio Lower Upper
113 100
55 148.6 154.1 194.1#
56 106.7 108.7 148.7#



#36
4-Chloro-3-methylphenol
Concen: 20.820 ng
RT: 8.710 min Scan# 1126
Delta R.T. -0.006 min
Lab File: BF125617.D
Acq: 24 Sep 2021 12:05

Tgt Ion:107 Resp: 247840
Ion Ratio Lower Upper
107 100
144 27.7 21.9 32.9
142 87.8 67.1 100.7









#37
2-Methylnaphthalene
Concen: 21.461 ng
RT: 8.881 min Scan# 1155
Delta R.T. -0.006 min
Lab File: BF125617.D
Acq: 24 Sep 2021 12:05




Instrument :
BNA_F
ClientSampleId :
SSTDIC020


Tgt Ion:142 Resp: 596854
Ion Ratio Lower Upper
142 100
141 89.6 71.2 106.8



#38
1-Methylnaphthalene
Concen: 21.353 ng
RT: 8.981 min Scan# 1172
Delta R.T. -0.006 min
Lab File: BF125617.D
Acq: 24 Sep 2021 12:05

Tgt Ion:142 Resp: 557205
Ion Ratio Lower Upper
142 100
141 94.3 76.0 114.0









#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.928 min Scan# 1333
Delta R.T. -0.000 min
Lab File: BF125617.D
Acq: 24 Sep 2021 12:05




Instrument :
BNA_F
ClientSampleId :
SSTDIC020


Tgt Ion:164	Resp:	425915
Ion Ratio	Lower	Upper
164	100	
162	102.9	81.3 121.9
160	44.3	38.0 57.0



#40
1,2,4,5-Tetrachlorobenzene
Concen: 20.757 ng
RT: 9.051 min Scan# 1184
Delta R.T. -0.000 min
Lab File: BF125617.D
Acq: 24 Sep 2021 12:05

Tgt Ion:216	Resp:	238047
Ion Ratio	Lower	Upper
216	100	
214	79.9	64.9 97.3
179	21.9	17.4 26.2
108	19.0	16.5 24.7









#41
Hexachlorocyclopentadiene
Concen: 15.424 ng
RT: 9.039 min Scan# 1182
Delta R.T. -0.000 min
Lab File: BF125617.D
Acq: 24 Sep 2021 12:05




Instrument :
BNA_F
ClientSampleId :
SSTDIC020


Tgt Ion:237 Resp: 98920
Ion Ratio Lower Upper
237 100
235 62.3 44.7 84.7
272 12.9 0.0 31.6



#42
2,4,6-Tribromophenol
Concen: 41.877 ng
RT: 10.710 min Scan# 1466
Delta R.T. -0.006 min
Lab File: BF125617.D
Acq: 24 Sep 2021 12:05

Tgt Ion:330 Resp: 174697
Ion Ratio Lower Upper
330 100
332 93.5 76.2 114.4
141 32.2 44.6 66.8#









#43
2,4,6-Trichlorophenol
Concen: 19.151 ng
RT: 9.157 min Scan# 1202
Delta R.T. -0.000 min
Lab File: BF125617.D
Acq: 24 Sep 2021 12:05




Instrument :
BNA_F
ClientSampleId :
SSTDIC020


Tgt Ion:196 Resp: 165473
Ion Ratio Lower Upper
196 100
198 96.6 74.9 112.3
200 30.0 27.0 40.4



#44
2,4,5-Trichlorophenol
Concen: 19.696 ng
RT: 9.192 min Scan# 1208
Delta R.T. -0.000 min
Lab File: BF125617.D
Acq: 24 Sep 2021 12:05

Tgt Ion:196 Resp: 178600
Ion Ratio Lower Upper
196 100
198 92.8 80.2 120.2
97 44.5 44.0 66.0
132 27.4 25.4 38.0









#45
2-Fluorobiphenyl
Concen: 42.023 ng
RT: 9.245 min Scan# 1217
Delta R.T. -0.006 min
Lab File: BF125617.D
Acq: 24 Sep 2021 12:05




Instrument :
BNA_F
ClientSampleId :
SSTDICC020


Tgt Ion:172 Resp: 1183856
Ion Ratio Lower Upper
172 100
171 35.6 26.5 39.7
170 23.9 18.8 28.2



#46
1,1'-Biphenyl
Concen: 21.040 ng
RT: 9.345 min Scan# 1234
Delta R.T. -0.006 min
Lab File: BF125617.D
Acq: 24 Sep 2021 12:05

Tgt Ion:154 Resp: 675209
Ion Ratio Lower Upper
154 100
153 42.0 20.0 60.0
76 12.2 0.0 30.3









#47
2-Chloronaphthalene
Concen: 20.986 ng
RT: 9.369 min Scan# 1238
Delta R.T. -0.006 min
Lab File: BF125617.D
Acq: 24 Sep 2021 12:05




Instrument :
BNA_F
ClientSampleId :
SSTDIC020


Tgt Ion:162 Resp: 559354
Ion Ratio Lower Upper
162 100
127 38.6 33.0 49.4
164 31.4 25.8 38.8



#48
2-Nitroaniline
Concen: 17.855 ng
RT: 9.463 min Scan# 1254
Delta R.T. -0.000 min
Lab File: BF125617.D
Acq: 24 Sep 2021 12:05

Tgt Ion: 65 Resp: 101553
Ion Ratio Lower Upper
65 100
92 75.1 55.5 83.3
138 134.4 93.5 140.3





#49

Acenaphthylene

Concen: 21.693 ng

RT: 9.786 min Scan# 1309

Delta R.T. -0.006 min

Lab File: BF125617.D

Acq: 24 Sep 2021 12:05

Instrument :

BNA_F

ClientSampleId :

SSTDIC020





Tgt Ion:152 Resp: 849749

Ion Ratio Lower Upper

152 100

151 20.2 15.7 23.5

153 12.9 9.8 14.8



#50

Dimethylphthalate

Concen: 20.769 ng

RT: 9.639 min Scan# 1284

Delta R.T. -0.006 min

Lab File: BF125617.D

Acq: 24 Sep 2021 12:05




Tgt Ion:163 Resp: 608799


Ion Ratio Lower Upper

163 100

194 3.4 2.6 3.8

164 9.8 7.4 11.0









#51
2,6-Dinitrotoluene
Concen: 16.692 ng
RT: 9.704 min Scan# 1295
Delta R.T. -0.000 min
Lab File: BF125617.D
Acq: 24 Sep 2021 12:05




Instrument :
BNA_F
ClientSampleId :
SSTDIC020


Tgt Ion:165 Resp: 97343
Ion Ratio Lower Upper
165 100
63 43.1 42.6 63.8
89 47.1 40.6 60.8



#52
Acenaphthene
Concen: 21.121 ng
RT: 9.957 min Scan# 1338
Delta R.T. -0.006 min
Lab File: BF125617.D
Acq: 24 Sep 2021 12:05

Tgt Ion:154 Resp: 522263
Ion Ratio Lower Upper
154 100
153 113.2 89.4 134.2
152 54.1 40.5 60.7









#53
3-Nitroaniline
Concen: 18.947 ng
RT: 9.869 min Scan# 1323
Delta R.T. -0.006 min
Lab File: BF125617.D
Acq: 24 Sep 2021 12:05




Instrument :
BNA_F
ClientSampleId :
SSTDIC020


Tgt Ion:138 Resp: 117356
Ion Ratio Lower Upper
138 100
108 15.3 9.3 13.9#
92 109.3 76.5 114.7



#54
2,4-Dinitrophenol
Concen: 13.884 ng
RT: 9.969 min Scan# 1340
Delta R.T. -0.006 min
Lab File: BF125617.D
Acq: 24 Sep 2021 12:05

Tgt Ion:184 Resp: 29171
Ion Ratio Lower Upper
184 100
63 73.0 59.3 88.9
154 280.3 68.6 102.8#









#55
Dibenzofuran
Concen: 21.823 ng
RT: 10.128 min Scan# 1367
Delta R.T. -0.006 min
Lab File: BF125617.D
Acq: 24 Sep 2021 12:05




Instrument :
BNA_F
ClientSampleId :
SSTDIC020


Tgt Ion:168 Resp: 800771
Ion Ratio Lower Upper
168 100
139 36.7 32.7 49.1
169 13.3 11.2 16.8



#56
4-Nitrophenol
Concen: 19.541 ng
RT: 10.016 min Scan# 1348
Delta R.T. -0.006 min
Lab File: BF125617.D
Acq: 24 Sep 2021 12:05

Tgt Ion:139 Resp: 98181
Ion Ratio Lower Upper
139 100
109 65.8 60.4 100.4
65 80.2 58.4 98.4









#57
2,4-Dinitrotoluene
Concen: 16.149 ng
RT: 10.104 min Scan# 1363
Delta R.T. -0.006 min
Lab File: BF125617.D
Acq: 24 Sep 2021 12:05




Instrument :
BNA_F
ClientSampleId :
SSTDIC020


Tgt Ion:165	Resp:	112861	
Ion Ratio	Lower	Upper	
165	100		
63	33.9	40.9	61.3#
89	65.8	59.7	89.5
182	2.6	0.0	0.0#



#58
Fluorene
Concen: 21.362 ng
RT: 10.475 min Scan# 1426
Delta R.T. -0.000 min
Lab File: BF125617.D
Acq: 24 Sep 2021 12:05

Tgt Ion:166	Resp:	602876	
Ion Ratio	Lower	Upper	
166	100		
165	98.6	78.1	117.1
167	12.5	11.1	16.7









#59
2,3,4,6-Tetrachlorophenol
Concen: 20.241 ng
RT: 10.245 min Scan# 1387
Delta R.T. -0.000 min
Lab File: BF125617.D
Acq: 24 Sep 2021 12:05




Instrument :
BNA_F
ClientSampleId :
SSTDIC020


Tgt Ion:232 Resp: 137854
Ion Ratio Lower Upper
232 100
131 47.5 45.4 68.0
130 2.3 0.0 0.0#
166 32.1 28.5 42.7



#60
Diethylphthalate
Concen: 20.358 ng
RT: 10.345 min Scan# 1404
Delta R.T. -0.000 min
Lab File: BF125617.D
Acq: 24 Sep 2021 12:05

Tgt Ion:149 Resp: 588131
Ion Ratio Lower Upper
149 100
177 21.9 16.5 24.7
150 11.8 10.1 15.1









#61
4-Chlorophenyl-phenylether
Concen: 21.100 ng
RT: 10.463 min Scan# 1424
Delta R.T. -0.006 min
Lab File: BF125617.D
Acq: 24 Sep 2021 12:05




Instrument :
BNA_F
ClientSampleId :
SSTDIC020


Tgt Ion:204 Resp: 264836
Ion Ratio Lower Upper
204 100
206 31.2 27.6 41.4
141 63.3 58.6 88.0



#62
4-Nitroaniline
Concen: 22.373 ng
RT: 10.480 min Scan# 1427
Delta R.T. -0.006 min
Lab File: BF125617.D
Acq: 24 Sep 2021 12:05

Tgt Ion:138 Resp: 125224
Ion Ratio Lower Upper
138 100
92 43.9 29.6 69.6
108 64.3 78.1 118.1#









#63
Azobenzene
Concen: 20.292 ng
RT: 10.622 min Scan# 1451
Delta R.T. -0.006 min
Lab File: BF125617.D
Acq: 24 Sep 2021 12:05




Instrument :
BNA_F
ClientSampleId :
SSTDIC020


Tgt Ion: 77	Resp: 569615
Ion Ratio	Lower Upper
77 100	
182 23.9	6.3 46.3
105 23.3	2.9 42.9
51 24.9	24.1 64.1



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.410 min Scan# 1585
Delta R.T. -0.000 min
Lab File: BF125617.D
Acq: 24 Sep 2021 12:05

Tgt Ion:188	Resp: 752334
Ion Ratio	Lower Upper
188 100	
94 9.7	6.7 10.1
80 9.8	7.5 11.3









#65
4,6-Dinitro-2-methylphenol
Concen: 12.914 ng
RT: 10.510 min Scan# 1432
Delta R.T. -0.006 min
Lab File: BF125617.D
Acq: 24 Sep 2021 12:05




Instrument :
BNA_F
ClientSampleId :
SSTDIC020


Tgt Ion:198 Resp: 47161
Ion Ratio Lower Upper
198 100
51 32.3 45.3 85.3#
105 41.3 30.6 70.6



#66
n-Nitrosodiphenylamine
Concen: 19.648 ng
RT: 10.580 min Scan# 1444
Delta R.T. -0.006 min
Lab File: BF125617.D
Acq: 24 Sep 2021 12:05

Tgt Ion:169 Resp: 536648
Ion Ratio Lower Upper
169 100
168 67.0 50.8 76.2
167 36.7 28.8 43.2









#67
4-Bromophenyl-phenylether
Concen: 18.745 ng
RT: 10.957 min Scan# 1508
Delta R.T. -0.000 min
Lab File: BF125617.D
Acq: 24 Sep 2021 12:05




Instrument :
BNA_F
ClientSampleId :
SSTDIC020


Tgt Ion:248 Resp: 161160
Ion Ratio Lower Upper
248 100
250 97.1 74.2 111.4
141 68.2 60.4 90.6



#68
Hexachlorobenzene
Concen: 19.954 ng
RT: 11.022 min Scan# 1519
Delta R.T. -0.000 min
Lab File: BF125617.D
Acq: 24 Sep 2021 12:05

Tgt Ion:284 Resp: 193344
Ion Ratio Lower Upper
284 100
142 31.0 36.4 54.6#
249 27.6 25.6 38.4





#69

Atrazine

Concen: 20.037 ng

RT: 11.104 min Scan# 1533

Delta R.T. -0.000 min

Lab File: BF125617.D

Acq: 24 Sep 2021 12:05

Tgt Ion:200 Resp: 147175

Ion Ratio Lower Upper

200 100

173 27.8 11.1 51.1





215 46.5 22.7 62.7

Instrument :

BNA_F

ClientSampleId :

SSTDIC020



#70

Pentachlorophenol

Concen: 18.139 ng

RT: 11.210 min Scan# 1551

Delta R.T. -0.000 min

Lab File: BF125617.D

Acq: 24 Sep 2021 12:05




Tgt Ion:266 Resp: 100203


Ion Ratio Lower Upper

266 100

268 62.2 48.9 73.3

264 62.1 55.9 83.9









#71
Phenanthrene
Concen: 21.594 ng
RT: 11.433 min Scan# 1589
Delta R.T. -0.000 min
Lab File: BF125617.D
Acq: 24 Sep 2021 12:05




Instrument :
BNA_F
ClientSampleId :
SSTDIC020


Tgt Ion:178 Resp: 912569
Ion Ratio Lower Upper
178 100
176 19.3 14.7 22.1
179 14.9 11.5 17.3



#72
Anthracene
Concen: 21.607 ng
RT: 11.486 min Scan# 1598
Delta R.T. -0.000 min
Lab File: BF125617.D
Acq: 24 Sep 2021 12:05

Tgt Ion:178 Resp: 910189
Ion Ratio Lower Upper
178 100
176 18.5 14.3 21.5
179 15.0 12.2 18.4









#73
Carbazole
Concen: 22.757 ng
RT: 11.633 min Scan# 1623
Delta R.T. -0.006 min
Lab File: BF125617.D
Acq: 24 Sep 2021 12:05




Instrument :
BNA_F
ClientSampleId :
SSTDIC020


Tgt Ion:167 Resp: 906656
Ion Ratio Lower Upper
167 100
166 21.0 17.4 26.0
139 12.6 11.0 16.4



#74
Di-n-butylphthalate
Concen: 19.427 ng
RT: 11.969 min Scan# 1680
Delta R.T. -0.000 min
Lab File: BF125617.D
Acq: 24 Sep 2021 12:05

Tgt Ion:149 Resp: 918595
Ion Ratio Lower Upper
149 100
150 8.9 7.4 11.2
104 4.4 4.1 6.1









#75
Fluoranthene
Concen: 23.065 ng
RT: 12.621 min Scan# 1791
Delta R.T. -0.000 min
Lab File: BF125617.D
Acq: 24 Sep 2021 12:05




Instrument :
BNA_F
ClientSampleId :
SSTDIC020


Tgt Ion:202 Resp: 956258
Ion Ratio Lower Upper
202 100
101 11.1 0.0 30.7
203 17.0 0.0 36.3



#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.045 min Scan# 2033
Delta R.T. -0.000 min
Lab File: BF125617.D
Acq: 24 Sep 2021 12:05

Tgt Ion:240 Resp: 504334
Ion Ratio Lower Upper
240 100
120 12.9 10.8 16.2
236 25.8 21.6 32.4





#77

Benzidine

Concen: 20.133 ng

RT: 12.739 min Scan# 1811

Delta R.T. -0.000 min

Lab File: BF125617.D

Acq: 24 Sep 2021 12:05

Instrument :





BNA_F

ClientSampleId :

SSTDIC020

Tgt Ion:184 Resp: 287906

Ion	Ratio	Lower	Upper
184	100		
185	13.9	10.8	16.2
183	11.2	8.1	12.1



#78

Pyrene

Concen: 20.630 ng

RT: 12.851 min Scan# 1830




Delta R.T. -0.000 min


Lab File: BF125617.D

Acq: 24 Sep 2021 12:05

Tgt Ion:202 Resp: 975008

Ion	Ratio	Lower	Upper
202	100		
200	21.2	15.3	22.9
203	17.0	13.4	20.0









#79
 Terphenyl-d14
 Concen: 41.043 ng
 RT: 12.992 min Scan# 1854
 Delta R.T. -0.000 min
 Lab File: BF125617.D
 Acq: 24 Sep 2021 12:05




Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC020


Tgt Ion:244 Resp: 1281845
 Ion Ratio Lower Upper
 244 100
 212 6.3 5.9 8.9
 122 11.8 9.0 13.4



#80
 Butylbenzylphthalate
 Concen: 16.785 ng
 RT: 13.468 min Scan# 1935
 Delta R.T. -0.000 min
 Lab File: BF125617.D
 Acq: 24 Sep 2021 12:05

Tgt Ion:149 Resp: 320551
 Ion Ratio Lower Upper
 149 100
 91 63.7 51.8 77.8
 206 20.4 14.5 21.7









#81
Benzo(a)anthracene
Concen: 20.932 ng
RT: 14.033 min Scan# 2031
Delta R.T. -0.006 min
Lab File: BF125617.D
Acq: 24 Sep 2021 12:05




Instrument :
BNA_F
ClientSampleId :
SSTDIC020


Tgt Ion:228 Resp: 718629
Ion Ratio Lower Upper
228 100
226 27.3 21.6 32.4
229 19.2 15.6 23.4



#82
3,3'-Dichlorobenzidine
Concen: 19.946 ng
RT: 13.998 min Scan# 2025
Delta R.T. -0.000 min
Lab File: BF125617.D
Acq: 24 Sep 2021 12:05

Tgt Ion:252 Resp: 209807
Ion Ratio Lower Upper
252 100
254 63.8 51.4 77.0
126 9.8 9.5 14.3









#83
Chrysene
Concen: 20.739 ng
RT: 14.074 min Scan# 2038
Delta R.T. -0.000 min
Lab File: BF125617.D
Acq: 24 Sep 2021 12:05




Instrument :
BNA_F
ClientSampleId :
SSTDIC020


Tgt Ion:228 Resp: 719224
Ion Ratio Lower Upper
228 100
226 29.5 22.6 34.0
229 18.5 15.9 23.9



#84
Bis(2-ethylhexyl)phthalate
Concen: 15.178 ng
RT: 14.027 min Scan# 2030
Delta R.T. -0.000 min
Lab File: BF125617.D
Acq: 24 Sep 2021 12:05

Tgt Ion:149 Resp: 354351
Ion Ratio Lower Upper
149 100
167 27.8 21.4 32.0
279 3.4 2.6 4.0









#85
Di-n-octyl phthalate
Concen: 11.763 ng
RT: 14.639 min Scan# 2134
Delta R.T. -0.000 min
Lab File: BF125617.D
Acq: 24 Sep 2021 12:05




Instrument :
BNA_F
ClientSampleId :
SSTDIC020


Tgt Ion:149 Resp: 456639
Ion Ratio Lower Upper
149 100
167 1.4 1.0 1.6
43 8.3 9.9 14.9#



#86
Perylene-d12
Concen: 20.000 ng
RT: 15.521 min Scan# 2284
Delta R.T. -0.000 min
Lab File: BF125617.D
Acq: 24 Sep 2021 12:05

Tgt Ion:264 Resp: 409409
Ion Ratio Lower Upper
264 100
260 23.4 19.6 29.4
265 20.5 17.2 25.8









#87
Indeno(1,2,3-cd)pyrene
Concen: 19.379 ng
RT: 16.998 min Scan# 2535
Delta R.T. -0.012 min
Lab File: BF125617.D
Acq: 24 Sep 2021 12:05




Instrument :
BNA_F
ClientSampleId :
SSTDIC020


Tgt Ion:276 Resp: 569558
Ion Ratio Lower Upper
276 100
138 28.8 24.1 36.1
277 24.5 18.8 28.2



#88
Benzo(b)fluoranthene
Concen: 19.498 ng
RT: 15.086 min Scan# 2210
Delta R.T. -0.000 min
Lab File: BF125617.D
Acq: 24 Sep 2021 12:05

Tgt Ion:252 Resp: 583818
Ion Ratio Lower Upper
252 100
253 21.5 16.8 25.2
125 10.5 9.8 14.8





#89

Benzo(k)fluoranthene

Concen: 21.320 ng

RT: 15.086 min Scan# 2210

Delta R.T. -0.035 min

Lab File: BF125617.D

Acq: 24 Sep 2021 12:05

Tgt Ion:252 Resp: 583818

Ion Ratio Lower Upper

252 100

253 21.5 17.8 26.8





125 10.5 7.2 10.8

Instrument :

BNA_F

ClientSampleId :

SSTDIC020



#90

Benzo(a)pyrene

Concen: 20.171 ng

RT: 15.457 min Scan# 2273

Delta R.T. -0.000 min

Lab File: BF125617.D

Acq: 24 Sep 2021 12:05




Tgt Ion:252 Resp: 525648


Ion Ratio Lower Upper

252 100

253 20.8 18.4 27.6

125 11.3 9.9 14.9









#91
Dibenzo(a,h)anthracene
Concen: 19.564 ng
RT: 17.015 min Scan# 2538
Delta R.T. -0.012 min
Lab File: BF125617.D
Acq: 24 Sep 2021 12:05

Instrument :
BNA_F
ClientSampleId :
SSTDIC020

Tgt Ion:278 Resp: 487319
Ion Ratio Lower Upper
278 100
139 18.7 14.9 22.3
279 22.8 19.7 29.5



#92
Benzo(g,h,i)perylene
Concen: 19.636 ng
RT: 17.445 min Scan# 2611
Delta R.T. -0.018 min
Lab File: BF125617.D
Acq: 24 Sep 2021 12:05

Tgt Ion:276 Resp: 482729
Ion Ratio Lower Upper
276 100
277 23.6 22.0 33.0
138 24.5 22.2 33.4

