

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF092421\
 Data File : BF125636.D
 Acq On : 24 Sep 2021 23:25
 Operator : JU/CG
 Sample : M3916-01
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 CONC-1

Quant Time: Sep 25 03:02:18 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF092421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 24 15:38:32 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.887	152	191325	20.00	ng	0.00
21) Naphthalene-d8	8.169	136	763093	20.00	ng	# 0.00
39) Acenaphthene-d10	9.922	164	394828	20.00	ng	0.00
64) Phenanthrene-d10	11.404	188	642990	20.00	ng	0.00
76) Chrysene-d12	14.039	240	354194	20.00	ng	0.00
86) Perylene-d12	15.515	264	398964	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.504	112	888072	76.47	ng	0.00
7) Phenol-d6	6.504	99	1194008	76.48	ng	0.00
23) Nitrobenzene-d5	7.445	82	750387	68.59	ng	0.00
42) 2,4,6-Tribromophenol	10.704	330	194715	52.89	ng	-0.01
45) 2-Fluorobiphenyl	9.245	172	1545426	67.59	ng	0.00
79) Terphenyl-d14	12.992	244	1255460	61.38	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF092421\
Data File : BF125636.D
Acq On : 24 Sep 2021 23:25
Operator : JU/CG
Sample : M3916-01
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
CONC-1

Quant Time: Sep 25 03:02:18 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF092421.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 24 15:38:32 2021
Response via : Initial Calibration

