

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF092421\
 Data File : BF125625.D
 Acq On : 24 Sep 2021 17:28
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 24 19:05:41 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF092421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 24 15:38:32 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	110	0.00
2	1,4-Dioxane	40.000	35.735	10.7	101	0.00
3	Pyridine	40.000	35.846	10.4	100	0.00
4	n-Nitrosodimethylamine	40.000	35.698	10.8	102	0.00
5 S	2-Fluorophenol	80.000	72.237	9.7	102	0.00
6	Aniline	40.000	35.205	12.0	102	0.00
7 S	Phenol-d6	80.000	71.001	11.2	102	0.00
8	2-Chlorophenol	40.000	36.757	8.1	103	0.00
9	Benzaldehyde	40.000	39.760	0.6	110	0.00
10 C	Phenol	40.000	35.735	10.7	100	0.00
11	bis(2-Chloroethyl)ether	40.000	34.939	12.7	103	0.00
12	1,3-Dichlorobenzene	40.000	35.458	11.4	104	0.00
13 C	1,4-Dichlorobenzene	40.000	34.928	12.7	102	0.00
14	1,2-Dichlorobenzene	40.000	35.125	12.2	103	0.00
15	Benzyl Alcohol	40.000	35.877	10.3	102	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	34.649	13.4	103	0.00
17	2-Methylphenol	40.000	35.554	11.1	103	0.00
18	Hexachloroethane	40.000	36.502	8.7	104	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	34.279	14.3	102	0.00
20	3+4-Methylphenols	40.000	35.397	11.5	102	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	110	0.00
22	Acetophenone	40.000	35.440	11.4	103	0.00
23 S	Nitrobenzene-d5	80.000	81.123	-1.4	111	0.00
24	Nitrobenzene	40.000	41.475	-3.7	109	0.00
25	Isophorone	40.000	35.425	11.4	103	0.00
26 C	2-Nitrophenol	40.000	41.144	-2.9	128	0.00
27	2,4-Dimethylphenol	40.000	36.757	8.1	105	0.00
28	bis(2-Chloroethoxy)methane	40.000	35.671	10.8	103	0.00
29 C	2,4-Dichlorophenol	40.000	37.949	5.1	104	0.00
30	1,2,4-Trichlorobenzene	40.000	35.692	10.8	102	0.00
31	Naphthalene	40.000	35.364	11.6	102	0.00
32	Benzoic acid	40.000	39.307	1.7	123	0.00
33	4-Chloroaniline	40.000	35.371	11.6	102	0.00
34 C	Hexachlorobutadiene	40.000	35.750	10.6	104	0.00
35	Caprolactam	40.000	36.940	7.7	104	0.00
36 C	4-Chloro-3-methylphenol	40.000	37.195	7.0	105	0.00
37	2-Methylnaphthalene	40.000	35.416	11.5	102	0.00
38	1-Methylnaphthalene	40.000	35.231	11.9	101	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	111	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	35.337	11.7	104	0.00
41 P	Hexachlorocyclopentadiene	40.000	38.654	3.4	105	0.00
42 S	2,4,6-Tribromophenol	80.000	78.053	2.4	107	0.00
43 C	2,4,6-Trichlorophenol	40.000	38.588	3.5	106	0.00
44	2,4,5-Trichlorophenol	40.000	38.396	4.0	105	0.00
45 S	2-Fluorobiphenyl	80.000	74.174	7.3	103	0.00
46	1,1'-Biphenyl	40.000	34.898	12.8	101	0.00
47	2-Chloronaphthalene	40.000	35.247	11.9	103	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF092421\
 Data File : BF125625.D
 Acq On : 24 Sep 2021 17:28
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 24 19:05:41 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF092421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 24 15:38:32 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	38.582	3.5	114	0.00
49	Acenaphthylene	40.000	35.337	11.7	102	0.00
50	Dimethylphthalate	40.000	35.464	11.3	104	0.00
51	2,6-Dinitrotoluene	40.000	40.189	-0.5	120	0.00
52 C	Acenaphthene	40.000	34.931	12.7	103	0.00
53	3-Nitroaniline	40.000	39.463	1.3	115	0.00
54 P	2,4-Dinitrophenol	40.000	43.244	-8.1	138	0.00
55	Dibenzofuran	40.000	34.992	12.5	103	0.00
56 P	4-Nitrophenol	40.000	42.146	-5.4	117	0.00
57	2,4-Dinitrotoluene	40.000	41.547	-3.9	127	0.00
58	Fluorene	40.000	32.974	17.6	102	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	39.050	2.4	107	0.00
60	Diethylphthalate	40.000	35.266	11.8	103	0.00
61	4-Chlorophenyl-phenylether	40.000	33.518	16.2	104	0.00
62	4-Nitroaniline	40.000	39.238	1.9	112	0.00
63	Azobenzene	40.000	35.203	12.0	104	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	113	0.00
65	4,6-Dinitro-2-methylphenol	40.000	40.165	-0.4	131	0.00
66 c	n-Nitrosodiphenylamine	40.000	34.786	13.0	103	0.00
67	4-Bromophenyl-phenylether	40.000	34.971	12.6	103	0.00
68	Hexachlorobenzene	40.000	34.799	13.0	103	0.00
69	Atrazine	40.000	37.124	7.2	104	0.00
70 C	Pentachlorophenol	40.000	41.182	-3.0	109	0.00
71	Phenanthrene	40.000	34.480	13.8	101	0.00
72	Anthracene	40.000	35.152	12.1	102	0.00
73	Carbazole	40.000	34.863	12.8	103	0.00
74	Di-n-butylphthalate	40.000	36.539	8.7	107	0.00
75 C	Fluoranthene	40.000	31.974	20.1#	103	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	116	0.00
77	Benzydine	40.000	36.412	9.0	100	0.00
78	Pyrene	40.000	37.863	5.3	103	0.00
79 S	Terphenyl-d14	80.000	73.515	8.1	105	0.00
80	Butylbenzylphthalate	40.000	39.462	1.3	112	0.00
81	Benzo(a)anthracene	40.000	35.645	10.9	107	0.00
82	3,3'-Dichlorobenzidine	40.000	36.425	8.9	114	0.00
83	Chrysene	40.000	35.209	12.0	108	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	37.401	6.5	116	0.00
85 c	Di-n-octyl phthalate	40.000	38.509	3.7	120	0.00
86 I	Perylene-d12	20.000	20.000	0.0	108	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	38.851	2.9	95	0.00
88	Benzo(b)fluoranthene	40.000	35.846	10.4	105	0.00
89	Benzo(k)fluoranthene	40.000	34.447	13.9	106	0.00
90 C	Benzo(a)pyrene	40.000	35.659	10.9	102	0.00
91	Dibenzo(a,h)anthracene	40.000	39.522	1.2	97	0.00
92	Benzo(g,h,i)perylene	40.000	39.462	1.3	95	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF092421\
Data File : BF125625.D
Acq On : 24 Sep 2021 17:28
Operator : JU/CG
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Sep 24 19:05:41 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF092421.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 24 15:38:32 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 1