

Data Path : Z:\HPCHEM1\BNA F\DATA\BF092517\
 Data File : BF098786.D
 Acq On : 25 Sep 2017 16:39
 Operator : SJ/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 26 06:48:13 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF092317.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Sep 23 13:50:58 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	0.00
2	1,4-Dioxane	0.600	0.595	0.8	98	-0.03
3	Pyridine	1.624	1.645	-1.3	104	-0.02
4	n-Nitrosodimethylamine	0.688	0.681	1.0	99	-0.02
5 S	2-Fluorophenol	1.256	1.257	-0.1	100	0.00
6	Aniline	2.092	2.087	0.2	101	0.00
7 S	Phenol-d6	1.550	1.560	-0.6	101	0.00
8	2-Chlorophenol	1.423	1.435	-0.8	103	0.00
9	Benzaldehyde	0.899	0.855	4.9	98	0.00
10 C	Phenol	1.733	1.734	-0.1	102	0.00
11	bis(2-Chloroethyl)ether	1.348	1.345	0.2	102	0.00
12	1,3-Dichlorobenzene	1.490	1.478	0.8	101	0.00
13 C	1,4-Dichlorobenzene	1.516	1.505	0.7	102	0.00
14	1,2-Dichlorobenzene	1.401	1.402	-0.1	102	0.00
15	Benzyl Alcohol	1.137	1.126	1.0	100	0.00
16	2,2'-oxybis(1-Chloropropane	2.099	2.066	1.6	102	0.00
17	2-Methylphenol	1.164	1.148	1.4	101	0.00
18	Hexachloroethane	0.545	0.541	0.7	102	0.00
19 P	n-Nitroso-di-n-propylamine	0.990	0.937	5.4	100	0.00
20	3+4-Methylphenols	1.473	1.445	1.9	99	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	102	0.00
22	Acetophenone	0.485	0.458	5.6	99	0.00
23 S	Nitrobenzene-d5	0.312	0.310	0.6	100	0.00
24	Nitrobenzene	0.354	0.345	2.5	100	0.00
25	Isophorone	0.645	0.625	3.1	102	0.00
26 C	2-Nitrophenol	0.170	0.182	-7.1	105	0.00
27	2,4-Dimethylphenol	0.321	0.310	3.4	100	0.00
28	bis(2-Chloroethoxy)methane	0.402	0.391	2.7	102	0.00
29 C	2,4-Dichlorophenol	0.276	0.274	0.7	102	0.00
30	1,2,4-Trichlorobenzene	0.276	0.272	1.4	102	0.00
31	Naphthalene	0.939	0.904	3.7	101	0.00
32	Benzoic acid	0.264	0.262	0.8	98	0.00
33	4-Chloroaniline	0.392	0.379	3.3	100	0.00
34 C	Hexachlorobutadiene	0.142	0.138	2.8	102	0.00
35	Caprolactam	0.088	0.085	3.4	101	0.00
36 C	4-Chloro-3-methylphenol	0.310	0.299	3.5	99	0.00
37	2-Methylnaphthalene	0.611	0.588	3.8	100	0.00
38 I	Acenaphthene-d10	1.000	1.000	0.0	101	0.00
39	1,2,4,5-Tetrachlorobenzene	0.596	0.584	2.0	100	0.00
40 P	Hexachlorocyclopentadiene	0.273	0.272	0.4	98	0.00
41 S	2,4,6-Tribromophenol	0.164	0.167	-1.8	99	0.00
42 C	2,4,6-Trichlorophenol	0.423	0.420	0.7	101	0.00
43	2,4,5-Trichlorophenol	0.422	0.423	-0.2	100	0.00
44 S	2-Fluorobiphenyl	1.198	1.200	-0.2	100	0.00

Data Path : Z:\HPCHEM1\BNA F\DATA\BF092517\
 Data File : BF098786.D
 Acq On : 25 Sep 2017 16:39
 Operator : SJ/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 26 06:48:13 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF092317.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Sep 23 13:50:58 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	1.719	1.681	2.2	100	0.00
46	2-Chloronaphthalene	1.291	1.264	2.1	100	0.00
47	2-Nitroaniline	0.395	0.402	-1.8	100	0.00
48	Acenaphthylene	1.976	1.919	2.9	100	0.00
49	Dimethylphthalate	1.509	1.485	1.6	102	0.00
50	2,6-Dinitrotoluene	0.329	0.334	-1.5	100	0.00
51 C	Acenaphthene	1.251	1.237	1.1	101	0.00
52	3-Nitroaniline	0.383	0.389	-1.6	98	0.00
53 P	2,4-Dinitrophenol	0.148	0.158	-6.8	103	0.00
54	Dibenzofuran	1.666	1.626	2.4	100	0.00
55 P	4-Nitrophenol	0.324	0.326	-0.6	99	0.00
56	2,4-Dinitrotoluene	0.426	0.443	-4.0	101	0.00
57	Fluorene	1.339	1.300	2.9	100	0.00
58	2,3,4,6-Tetrachlorophenol	0.291	0.285	2.1	98	0.00
59	Diethylphthalate	1.475	1.428	3.2	99	0.00
60	4-Chlorophenyl-phenylether	0.602	0.583	3.2	100	0.00
61	4-Nitroaniline	0.370	0.378	-2.2	100	0.00
62	Azobenzene	1.356	1.311	3.3	99	0.00
63 I	Phenanthrene-d10	1.000	1.000	0.0	97	0.00
64	4,6-Dinitro-2-methylphenol	0.122	0.129	-5.7	102	0.00
65 c	n-Nitrosodiphenylamine	0.693	0.678	2.2	99	0.00
66	4-Bromophenyl-phenylether	0.196	0.197	-0.5	100	0.00
67	Hexachlorobenzene	0.209	0.208	0.5	100	0.00
68	Atrazine	0.208	0.198	4.8	94	0.00
69 C	Pentachlorophenol	0.130	0.130	0.0	94	0.00
70	Phenanthrene	1.071	1.045	2.4	99	0.00
71	Anthracene	1.095	1.071	2.2	98	0.00
72	Carbazole	1.073	1.043	2.8	97	0.00
73	Di-n-butylphthalate	1.291	1.279	0.9	98	0.00
74 C	Fluoranthene	1.084	1.051	3.0	98	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	98	0.00
76	Benzidine	0.740	0.729	1.5	98	0.00
77	Pyrene	1.525	1.484	2.7	98	0.00
78 S	Terphenyl-d14	0.879	0.871	0.9	99	0.00
79	Butylbenzylphthalate	0.717	0.716	0.1	97	0.00
80	Benzo(a)anthracene	1.211	1.164	3.9	98	0.00
81	3,3'-Dichlorobenzidine	0.444	0.446	-0.5	100	0.00
82	Chrysene	1.153	1.117	3.1	98	0.00
83	Bis(2-ethylhexyl)phthalate	0.987	1.004	-1.7	100	0.00
84 c	Di-n-octyl phthalate	1.589	1.638	-3.1	105	0.00
85	Indeno(1,2,3-cd)pyrene	0.922	0.979	-6.2	115	0.00
86 I	Perylene-d12	1.000	1.000	0.0	106	0.00
87	Benzo(b)fluoranthene	1.230	1.217	1.1	103	0.00

Data Path : Z:\HPCHEM1\BNA F\DATA\BF092517\
Data File : BF098786.D
Acq On : 25 Sep 2017 16:39
Operator : SJ/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Sep 26 06:48:13 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF092317.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Sep 23 13:50:58 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.215	1.147	5.6	102	0.00
89 C	Benzo(a)pyrene	1.129	1.113	1.4	106	0.00
90	Dibenzo(a,h)anthracene	0.903	0.944	-4.5	115	0.00
91	Benzo(a,h,i)perylene	0.893	0.943	-5.6	116	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0