



Data Path : Z:\svoasrv\HPCHEM1\BNA F\Data\BF092518\
Data File : BF109307.D
Acq On : 26 Sep 2018 1:13
Operator : JU/SJ
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Oct 01 08:42:40 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF092218.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Sep 22 13:32:18 2018
Response via : Initial Calibration



TIC: BF109307.D

(76) Benzidine



12.545min (+0.000) 0.00ng

response 409352

Ion	Exp%	Act%
184.00	100	100
185.00	15.70	14.51
183.00	11.60	11.57
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA F\Data\BF092518\
Data File : BF109307.D
Acq On : 26 Sep 2018 1:13
Operator : JU/SJ
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Oct 01 08:42:40 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF092218.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Sep 22 13:32:18 2018
Response via : Initial Calibration



TIC: BF109307.D

(69) Pentachlorophenol (C)

11.028min (+0.000) 58543.53ng

response 49589

Ion	Exp%	Act%
265.70	100	100
268.00	63.90	61.12
264.00	61.90	62.69
0.00	0.00	0.00