


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF092518\
 Data File : BF109307.D
 Acq On : 26 Sep 2018 1:13
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF092218.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Sat Sep 22 13:32:18 2018



AutoFind: Scans 1597, 1598, 1599; Background Corrected with Scan 1593

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	36.5	26389	PASS
68	69	0.00	2	1.7	470	PASS
69	198	0.00	100	37.8	27362	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	49.2	35610	PASS
197	198	0.00	2	0.6	460	PASS
198	198	100	100	100.0	72360	PASS
199	198	5	9	6.6	4789	PASS
275	198	10	60	24.8	17967	PASS
365	198	1	100	3.0	2137	PASS
441	198	0.01	100	11.5	8336	PASS
442	198	50	100	71.8	51961	PASS
443	442	15	24	18.4	9566	PASS