

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF092615\
 Data File : BF081839.D
 Acq On : 26 Sep 2015 20:58
 Operator : UM/IZ
 Sample : G3754-11
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 EQUIPMENTBLANK

Quant Time: Sep 28 07:59:01 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF092415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 25 06:02:49 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.94	152	121492	20.00	ng	-0.01
21) Naphthalene-d8	8.23	136	454942	20.00	ng	-0.01
38) Acenaphthene-d10	9.98	164	203082	20.00	ng	-0.02
63) Phenanthrene-d10	11.46	188	349487	20.00	ng	-0.01
75) Chrysene-d12	14.12	240	289051	20.00	ng	0.00
86) Perylene-d12	15.58	264	287137	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.53	112	387207	53.92	ng	-0.01
7) Phenol-d6	6.56	99	313451	34.51	ng	-0.01
23) Nitrobenzene-d5	7.51	82	686964	99.63	ng	-0.01
41) 2,4,6-Tribromophenol	10.78	330	257561	146.08	ng	-0.01
44) 2-Fluorobiphenyl	9.30	172	1100858	88.98	ng	-0.01
78) Terphenyl-d14	13.06	244	949145	75.01	ng	0.00

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF092615\
Data File : BF081839.D
Acq On : 26 Sep 2015 20:58
Operator : UM/IZ
Sample : G3754-11
Misc :
ALS Vial : 53 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
EQUIPMENTBLANK

Quant Time: Sep 28 07:59:01 2015
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF092415.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 25 06:02:49 2015
Response via : Initial Calibration

