


Data Path : Z:\HPCHEM1\BNA F\DATA\BF092616\
 Data File : BF090215.D
 Acq On : 26 Sep 2016 11:57
 Operator : UM/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF092016.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Sep 26 15:58:38 2016



AutoFind: Scans 846, 847, 848; Background Corrected with Scan 843

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	33.8	13080	PASS
68	69	0.00	2	1.2	177	PASS
69	198	0.00	100	36.7	14196	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	52.2	20195	PASS
197	198	0.00	2	0.6	218	PASS
198	198	100	100	100.0	38693	PASS
199	198	5	9	6.2	2400	PASS
275	198	10	60	25.4	9841	PASS
365	198	1	100	3.0	1180	PASS
441	442	0.01	24	15.4	5326	PASS
442	198	50	100	89.3	34554	PASS
443	442	15	24	20.0	6924	PASS