

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF092618\
 Data File : BF109337.D
 Acq On : 26 Sep 2018 15:35
 Operator : JU/SJ
 Sample : J5056-01
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 RA-BASELINE-WATER

Quant Time: Sep 26 16:46:15 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF092218.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Sep 22 13:32:18 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.70	152	107952	20.00	ng	0.00
21) Naphthalene-d8	7.99	136	397559	20.00	ng	0.00
38) Acenaphthene-d10	9.74	164	163011	20.00	ng	0.00
63) Phenanthrene-d10	11.22	188	253265	20.00	ng	0.00
75) Chrysene-d12	13.89	240	137448	20.00	ng	0.04
86) Perylene-d12	15.34	264	113727	20.00	ng	0.11
System Monitoring Compounds						
5) 2-Fluorophenol	5.31	112	418591	63.87	ng	0.00
7) Phenol-d6	6.34	99	393576	47.06	ng	-0.02
23) Nitrobenzene-d5	7.27	82	548492	81.44	ng	-0.01
41) 2,4,6-Tribromophenol	10.53	330	175487	109.21	ng	0.00
44) 2-Fluorobiphenyl	9.07	172	930585	86.10	ng	0.00
78) Terphenyl-d14	12.82	244	598790	106.92	ng	0.01
Target Compounds						
73) Di-n-butylphthalate	11.79	149	57759	3.932	ng	Qvalue # 80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF092618\
Data File : BF109337.D
Acq On : 26 Sep 2018 15:35
Operator : JU/SJ
Sample : J5056-01
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
RA-BASELINE-WATER

Quant Time: Sep 26 16:46:15 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF092218.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Sep 22 13:32:18 2018
Response via : Initial Calibration

