




Data Path : Z:\svoasrv\HPCHEM1\BNA F\Data\BF092618\
Data File : BF109330.D
Acq On : 26 Sep 2018 12:27
Operator : JU/SJ
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Sep 26 18:16:43 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF092218.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Sep 22 13:32:18 2018
Response via : Initial Calibration



(69) Pentachlorophenol (C)




11.027min (+0.000) 49666.01ng

response 117093

Ion	Exp%	Act%
265.70	100	100
268.00	63.90	63.42
264.00	61.90	64.01
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA F\Data\BF092618\
Data File : BF109330.D
Acq On : 26 Sep 2018 12:27
Operator : JU/SJ
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Sep 26 18:16:43 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF092218.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Sep 22 13:32:18 2018
Response via : Initial Calibration



(76) Benzidine

12.545min (+0.000) 0.00ng

response 658953

Ion	Exp%	Act%
184.00	100	100
185.00	15.70	14.40
183.00	11.60	11.82
0.00	0.00	0.00