


Data Path : Z:\svoasrv\HPCHEM1\BNA\_F\Data\BF092618\  
 Data File : BF109330.D  
 Acq On : 26 Sep 2018 12:27  
 Operator : JU/SJ  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA\_F\METHODS\8270-BF092218.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Sat Sep 22 13:32:18 2018



AutoFind: Scans 1597, 1598, 1599; Background Corrected with Scan 1592

| Target<br>Mass | Rel. to<br>Mass | Lower<br>Limit% | Upper<br>Limit% | Rel.<br>Abn% | Raw<br>Abn | Result<br>Pass/Fail |
|----------------|-----------------|-----------------|-----------------|--------------|------------|---------------------|
| 51             | 198             | 10              | 80              | 35.9         | 42114      | PASS                |
| 68             | 69              | 0.00            | 2               | 1.6          | 699        | PASS                |
| 69             | 198             | 0.00            | 100             | 37.3         | 43778      | PASS                |
| 70             | 69              | 0.00            | 2               | 0.0          | 0          | PASS                |
| 127            | 198             | 10              | 80              | 48.8         | 57224      | PASS                |
| 197            | 198             | 0.00            | 2               | 0.7          | 865        | PASS                |
| 198            | 198             | 100             | 100             | 100.0        | 117333     | PASS                |
| 199            | 198             | 5               | 9               | 6.6          | 7687       | PASS                |
| 275            | 198             | 10              | 60              | 24.6         | 28818      | PASS                |
| 365            | 198             | 1               | 100             | 2.9          | 3363       | PASS                |
| 441            | 198             | 0.01            | 100             | 11.8         | 13802      | PASS                |
| 442            | 198             | 50              | 100             | 73.0         | 85618      | PASS                |
| 443            | 442             | 15              | 24              | 18.2         | 15553      | PASS                |