


Data Path : Z:\HPCHEM1\BNA_F\DATA\BF092717\
 Data File : BF098851.D
 Acq On : 27 Sep 2017 9:28
 Operator : SJ/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF092317.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Sat Sep 23 13:50:58 2017



AutoFind: Scans 1631, 1632, 1633; Background Corrected with Scan 1626

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	35.4	48733	PASS
68	69	0.00	2	1.9	863	PASS
69	198	0.00	100	32.7	45016	PASS
70	69	0.00	2	0.6	275	PASS
127	198	10	80	48.7	67051	PASS
197	198	0.00	2	0.5	732	PASS
198	198	100	100	100.0	137757	PASS
199	198	5	9	6.4	8791	PASS
275	198	10	60	23.0	31623	PASS
365	198	1	100	2.6	3645	PASS
441	442	0.01	24	15.4	19211	PASS
442	198	50	100	90.7	124914	PASS
443	442	15	24	18.5	23121	PASS