


Data Path : Z:\HPCHEM1\BNA_F\DATA\BF092717\
 Data File : BF098891.D
 Acq On : 28 Sep 2017 5:01
 Operator : SJ/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF092317.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Sat Sep 23 13:50:58 2017



AutoFind: Scans 1631, 1632, 1633; Background Corrected with Scan 1626

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	38.7	41324	PASS
68	69	0.00	2	1.5	554	PASS
69	198	0.00	100	35.5	37827	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	49.8	53069	PASS
197	198	0.00	2	0.9	951	PASS
198	198	100	100	100.0	106661	PASS
199	198	5	9	6.4	6867	PASS
275	198	10	60	23.2	24724	PASS
365	198	1	100	2.6	2783	PASS
441	442	0.01	24	15.6	13027	PASS
442	198	50	100	78.3	83469	PASS
443	442	15	24	19.5	16271	PASS