

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF092722\
 Data File : BF130422.D
 Acq On : 27 Sep 2022 13:19
 Operator : CG\JU
 Sample : PB147890BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB147890BL

Quant Time: Sep 28 05:10:13 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 28 02:59:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.640	152	99767	20.000	ng	0.00
21) Naphthalene-d8	7.922	136	405169	20.000	ng	0.00
39) Acenaphthene-d10	9.675	164	221081	20.000	ng	0.00
64) Phenanthrene-d10	11.157	188	373030	20.000	ng	0.00
76) Chrysene-d12	13.786	240	230018	20.000	ng	0.00
86) Perylene-d12	15.168	264	180660	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.257	112	836788	145.477	ng	0.01
7) Phenol-d6	6.287	99	1023985	142.277	ng	0.00
23) Nitrobenzene-d5	7.210	82	656933	82.834	ng	0.00
42) 2,4,6-Tribromophenol	10.463	330	331472	156.474	ng	0.00
45) 2-Fluorobiphenyl	9.004	172	1311630	86.393	ng	0.00
79) Terphenyl-d14	12.745	244	1410412	101.572	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF092722\
Data File : BF130422.D
Acq On : 27 Sep 2022 13:19
Operator : CG\JU
Sample : PB147890BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB147890BL

Quant Time: Sep 28 05:10:13 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091422.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 28 02:59:31 2022
Response via : Initial Calibration

