

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF092722\
 Data File : BF130421.D
 Acq On : 27 Sep 2022 12:10
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 27 18:39:59 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 27 01:00:47 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	106	0.00
2	1,4-Dioxane	40.000	34.588	13.5	91	0.00
3	Pyridine	40.000	36.250	9.4	99	0.00
4	n-Nitrosodimethylamine	40.000	32.338	19.2	85	0.00
5 S	2-Fluorophenol	80.000	78.786	1.5	106	0.00
6	Aniline	40.000	37.346	6.6	100	0.00
7 S	Phenol-d6	80.000	77.162	3.5	103	0.00
8	2-Chlorophenol	40.000	40.263	-0.7	108	0.00
9	Benzaldehyde	40.000	34.865	12.8	95	0.00
10 C	Phenol	40.000	37.974	5.1	101	0.00
11	bis(2-Chloroethyl)ether	40.000	37.903	5.2	101	0.00
12	1,3-Dichlorobenzene	40.000	39.949	0.1	108	0.00
13 C	1,4-Dichlorobenzene	40.000	39.913	0.2	108	0.00
14	1,2-Dichlorobenzene	40.000	39.855	0.4	106	0.00
15	Benzyl Alcohol	40.000	31.116	22.2	80	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	32.132	19.7	86	0.00
17	2-Methylphenol	40.000	38.638	3.4	102	0.00
18	Hexachloroethane	40.000	38.430	3.9	102	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	35.035	12.4	92	0.00
20	3+4-Methylphenols	40.000	38.872	2.8	103	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	102	0.00
22	Acetophenone	40.000	38.935	2.7	99	0.00
23 S	Nitrobenzene-d5	80.000	77.088	3.6	97	0.00
24	Nitrobenzene	40.000	37.739	5.7	96	0.00
25	Isophorone	40.000	37.849	5.4	96	0.00
26 C	2-Nitrophenol	40.000	45.753	-14.4	111	0.00
27	2,4-Dimethylphenol	40.000	39.523	1.2	100	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.150	2.1	101	0.00
29 C	2,4-Dichlorophenol	40.000	42.859	-7.1	108	0.00
30	1,2,4-Trichlorobenzene	40.000	42.718	-6.8	109	0.00
31	Naphthalene	40.000	40.798	-2.0	105	0.00
32	Benzoic acid	40.000	39.044	2.4	92	0.00
33	4-Chloroaniline	40.000	41.692	-4.2	106	0.00
34 C	Hexachlorobutadiene	40.000	42.752	-6.9	109	0.00
35	Caprolactam	40.000	41.872	-4.7	103	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.238	-0.6	101	0.00
37	2-Methylnaphthalene	40.000	41.016	-2.5	105	0.00
38	1-Methylnaphthalene	40.000	40.898	-2.2	105	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	103	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	42.233	-5.6	112	0.00
41 P	Hexachlorocyclopentadiene	40.000	34.780	13.0	87	0.00
42 S	2,4,6-Tribromophenol	80.000	89.023	-11.3	113	0.00
43 C	2,4,6-Trichlorophenol	40.000	42.104	-5.3	107	0.00
44	2,4,5-Trichlorophenol	40.000	42.543	-6.4	111	0.00
45 S	2-Fluorobiphenyl	80.000	80.773	-1.0	107	0.00
46	1,1'-Biphenyl	40.000	40.936	-2.3	107	0.00
47	2-Chloronaphthalene	40.000	40.645	-1.6	107	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF092722\
 Data File : BF130421.D
 Acq On : 27 Sep 2022 12:10
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 27 18:39:59 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 27 01:00:47 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	37.231	6.9	93	0.00
49	Acenaphthylene	40.000	40.845	-2.1	107	0.00
50	Dimethylphthalate	40.000	40.846	-2.1	106	0.00
51	2,6-Dinitrotoluene	40.000	43.299	-8.2	112	0.00
52 C	Acenaphthene	40.000	40.719	-1.8	107	0.00
53	3-Nitroaniline	40.000	42.097	-5.2	107	0.00
54 P	2,4-Dinitrophenol	40.000	39.436	1.4	105	0.00
55	Dibenzofuran	40.000	40.919	-2.3	108	0.00
56 P	4-Nitrophenol	40.000	37.835	5.4	92	0.00
57	2,4-Dinitrotoluene	40.000	43.191	-8.0	108	0.00
58	Fluorene	40.000	40.815	-2.0	105	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	42.775	-6.9	112	0.00
60	Diethylphthalate	40.000	39.798	0.5	103	0.00
61	4-Chlorophenyl-phenylether	40.000	41.769	-4.4	109	0.00
62	4-Nitroaniline	40.000	41.356	-3.4	104	0.00
63	Azobenzene	40.000	35.742	10.6	92	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	105	0.00
65	4,6-Dinitro-2-methylphenol	40.000	46.105	-15.3	117	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.136	-0.3	107	0.00
67	4-Bromophenyl-phenylether	40.000	42.823	-7.1	115	0.00
68	Hexachlorobenzene	40.000	43.495	-8.7	116	0.00
69	Atrazine	40.000	40.567	-1.4	107	0.00
70 C	Pentachlorophenol	40.000	39.187	2.0	97	0.00
71	Phenanthrene	40.000	40.175	-0.4	108	0.00
72	Anthracene	40.000	40.743	-1.9	108	0.00
73	Carbazole	40.000	40.119	-0.3	106	0.00
74	Di-n-butylphthalate	40.000	39.988	0.0	104	0.00
75 C	Fluoranthene	40.000	41.344	-3.4	108	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	102	0.00
77	Benzydine	40.000	36.624	8.4	88	0.00
78	Pyrene	40.000	42.725	-6.8	109	0.00
79 S	Terphenyl-d14	80.000	84.988	-6.2	108	0.00
80	Butylbenzylphthalate	40.000	42.749	-6.9	106	0.00
81	Benzo(a)anthracene	40.000	40.974	-2.4	106	0.00
82	3,3'-Dichlorobenzidine	40.000	44.243	-10.6	112	0.00
83	Chrysene	40.000	40.747	-1.9	107	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	40.824	-2.1	102	0.00
85 c	Di-n-octyl phthalate	40.000	40.280	-0.7	105	0.00
86 I	Perylene-d12	20.000	20.000	0.0	111	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	41.841	-4.6	112	0.00
88	Benzo(b)fluoranthene	40.000	39.238	1.9	109	0.00
89	Benzo(k)fluoranthene	40.000	41.209	-3.0	115	0.00
90 C	Benzo(a)pyrene	40.000	41.655	-4.1	115	0.00
91	Dibenzo(a,h)anthracene	40.000	42.161	-5.4	113	0.00
92	Benzo(g,h,i)perylene	40.000	41.411	-3.5	111	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF092722\
Data File : BF130421.D
Acq On : 27 Sep 2022 12:10
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Sep 27 18:39:59 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091422.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 27 01:00:47 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0