


Data Path : Z:\HPCHEM1\BNA_F\DATA\BF092815\
 Data File : BF081860.D
 Acq On : 29 Sep 2015 1:58
 Operator : UM/IZ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF092815.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Sep 28 18:10:42 2015



AutoFind: Scans 837, 838, 839; Background Corrected with Scan 834

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	33.0	9080	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	29.1	7999	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	44.2	12163	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	27492	PASS
199	198	5	9	5.8	1608	PASS
275	198	10	60	26.8	7379	PASS
365	198	1	100	3.1	859	PASS
441	442	0.01	24	15.2	4098	PASS
442	198	50	100	97.8	26899	PASS
443	442	15	24	19.0	5110	PASS