


Data Path : Z:\HPCHEM1\BNA_F\DATA\BF092816\
 Data File : BF090282.D
 Acq On : 28 Sep 2016 16:53
 Operator : UM/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF092016.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Sep 26 15:58:38 2016



AutoFind: Scans 846, 847, 848; Background Corrected with Scan 842

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	32.1	17005	PASS
68	69	0.00	2	1.2	231	PASS
69	198	0.00	100	36.8	19454	PASS
70	69	0.00	2	0.5	103	PASS
127	198	10	80	52.3	27688	PASS
197	198	0.00	2	0.5	278	PASS
198	198	100	100	100.0	52928	PASS
199	198	5	9	6.7	3545	PASS
275	198	10	60	23.7	12519	PASS
365	198	1	100	2.9	1514	PASS
441	442	0.01	24	15.7	7072	PASS
442	198	50	100	84.9	44944	PASS
443	442	15	24	19.3	8689	PASS