

Data Path : Z:\HPCHEM1\BNA F\Data\BF092817\
 Data File : BF098913.D
 Acq On : 28 Sep 2017 18:07
 Operator : SJ/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 29 01:41:03 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF092317.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Sep 23 13:50:58 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
2	1,4-Dioxane	0.600	0.582	3.0	95	-0.04
3	Pyridine	1.624	1.636	-0.7	102	-0.02
4	n-Nitrosodimethylamine	0.688	0.683	0.7	99	-0.02
5 S	2-Fluorophenol	1.256	1.241	1.2	98	0.00
6	Aniline	2.092	2.056	1.7	99	0.00
7 S	Phenol-d6	1.550	1.541	0.6	99	0.00
8	2-Chlorophenol	1.423	1.398	1.8	99	0.00
9	Benzaldehyde	0.899	0.835	7.1	95	0.00
10 C	Phenol	1.733	1.725	0.5	100	0.00
11	bis(2-Chloroethyl)ether	1.348	1.333	1.1	100	0.00
12	1,3-Dichlorobenzene	1.490	1.462	1.9	99	0.00
13 C	1,4-Dichlorobenzene	1.516	1.485	2.0	99	0.00
14	1,2-Dichlorobenzene	1.401	1.370	2.2	98	0.00
15	Benzyl Alcohol	1.137	1.129	0.7	99	0.00
16	2,2'-oxybis(1-Chloropropane	2.099	2.076	1.1	101	0.00
17	2-Methylphenol	1.164	1.147	1.5	99	0.00
18	Hexachloroethane	0.545	0.533	2.2	99	-0.01
19 P	n-Nitroso-di-n-propylamine	0.990	0.926	6.5	98	0.00
20	3+4-Methylphenols	1.473	1.400	5.0	95	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	100	0.00
22	Acetophenone	0.485	0.451	7.0	97	0.00
23 S	Nitrobenzene-d5	0.312	0.314	-0.6	100	0.00
24	Nitrobenzene	0.354	0.350	1.1	100	0.00
25	Isophorone	0.645	0.632	2.0	101	0.00
26 C	2-Nitrophenol	0.170	0.183	-7.6	105	0.00
27	2,4-Dimethylphenol	0.321	0.308	4.0	98	0.00
28	bis(2-Chloroethoxy)methane	0.402	0.393	2.2	101	0.00
29 C	2,4-Dichlorophenol	0.276	0.272	1.4	100	0.00
30	1,2,4-Trichlorobenzene	0.276	0.269	2.5	99	0.00
31	Naphthalene	0.939	0.906	3.5	100	0.00
32	Benzoic acid	0.264	0.217	17.8	80	0.00
33	4-Chloroaniline	0.392	0.388	1.0	101	0.00
34 C	Hexachlorobutadiene	0.142	0.136	4.2	99	0.00
35	Caprolactam	0.088	0.086	2.3	101	0.00
36 C	4-Chloro-3-methylphenol	0.310	0.303	2.3	99	0.00
37	2-Methylnaphthalene	0.611	0.589	3.6	99	0.00
38 I	Acenaphthene-d10	1.000	1.000	0.0	102	0.00
39	1,2,4,5-Tetrachlorobenzene	0.596	0.575	3.5	100	0.00
40 P	Hexachlorocyclopentadiene	0.273	0.246	9.9	90	0.00
41 S	2,4,6-Tribromophenol	0.164	0.162	1.2	98	0.00
42 C	2,4,6-Trichlorophenol	0.423	0.408	3.5	99	0.00
43	2,4,5-Trichlorophenol	0.422	0.421	0.2	101	0.00
44 S	2-Fluorobiphenyl	1.198	1.169	2.4	99	0.00

Data Path : Z:\HPCHEM1\BNA F\Data\BF092817\
 Data File : BF098913.D
 Acq On : 28 Sep 2017 18:07
 Operator : SJ/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 29 01:41:03 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF092317.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Sep 23 13:50:58 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	1.719	1.625	5.5	98	0.00
46	2-Chloronaphthalene	1.291	1.242	3.8	100	0.00
47	2-Nitroaniline	0.395	0.406	-2.8	102	0.00
48	Acenaphthylene	1.976	1.894	4.1	99	0.00
49	Dimethylphthalate	1.509	1.474	2.3	102	0.00
50	2,6-Dinitrotoluene	0.329	0.336	-2.1	102	0.00
51 C	Acenaphthene	1.251	1.224	2.2	101	0.00
52	3-Nitroaniline	0.383	0.399	-4.2	102	0.00
53 P	2,4-Dinitrophenol	0.148	0.159	-7.4	104	0.00
54	Dibenzofuran	1.666	1.594	4.3	99	0.00
55 P	4-Nitrophenol	0.324	0.319	1.5	97	0.00
56	2,4-Dinitrotoluene	0.426	0.446	-4.7	103	0.00
57	Fluorene	1.339	1.292	3.5	101	0.00
58	2,3,4,6-Tetrachlorophenol	0.291	0.284	2.4	99	0.00
59	Diethylphthalate	1.475	1.420	3.7	100	0.00
60	4-Chlorophenyl-phenylether	0.602	0.580	3.7	100	0.00
61	4-Nitroaniline	0.370	0.382	-3.2	102	0.00
62	Azobenzene	1.356	1.311	3.3	100	0.00
63 I	Phenanthrene-d10	1.000	1.000	0.0	99	0.00
64	4,6-Dinitro-2-methylphenol	0.122	0.129	-5.7	105	0.00
65 c	n-Nitrosodiphenylamine	0.693	0.669	3.5	100	0.00
66	4-Bromophenyl-phenylether	0.196	0.190	3.1	99	0.00
67	Hexachlorobenzene	0.209	0.204	2.4	100	0.00
68	Atrazine	0.208	0.200	3.8	97	0.00
69 C	Pentachlorophenol	0.130	0.125	3.8	93	0.00
70	Phenanthrene	1.071	1.018	4.9	99	0.00
71	Anthracene	1.095	1.055	3.7	99	0.00
72	Carbazole	1.073	1.037	3.4	99	0.00
73	Di-n-butylphthalate	1.291	1.254	2.9	99	0.00
74 C	Fluoranthene	1.084	1.029	5.1	98	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	97	0.00
76	Benzidine	0.740	0.717	3.1	96	0.00
77	Pyrene	1.525	1.499	1.7	98	0.00
78 S	Terphenyl-d14	0.879	0.880	-0.1	99	0.00
79	Butylbenzylphthalate	0.717	0.758	-5.7	102	0.00
80	Benzo(a)anthracene	1.211	1.190	1.7	99	0.00
81	3,3'-Dichlorobenzidine	0.444	0.429	3.4	95	0.00
82	Chrysene	1.153	1.121	2.8	97	0.00
83	Bis(2-ethylhexyl)phthalate	0.987	1.038	-5.2	103	0.00
84 c	Di-n-octyl phthalate	1.589	1.706	-7.4	108	0.00
85	Indeno(1,2,3-cd)pyrene	0.922	0.877	4.9	102	0.01
86 I	Perylene-d12	1.000	1.000	0.0	101	0.00
87	Benzo(b)fluoranthene	1.230	1.293	-5.1	104	0.00

Data Path : Z:\HPCHEM1\BNA F\Data\BF092817\
Data File : BF098913.D
Acq On : 28 Sep 2017 18:07
Operator : SJ/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Sep 29 01:41:03 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF092317.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Sep 23 13:50:58 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.215	1.114	8.3	94	0.00
89 C	Benzo(a)pyrene	1.129	1.107	1.9	100	0.00
90	Dibenzo(a,h)anthracene	0.903	0.904	-0.1	104	0.01
91	Benzo(a,h,i)perylene	0.893	0.879	1.6	102	0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0