

Data Path : Z:\HPCHEM1\BNA F\DATA\BF092915\
 Data File : BF081903.D
 Acq On : 30 Sep 2015 8:02
 Operator : UM/IZ
 Sample : G3754-02DL2 25X
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 MW-2BDL2

Quant Time: Sep 30 10:55:10 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF092815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Sep 28 18:10:42 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.91	152	115161	20.00	ng	-0.02
21) Naphthalene-d8	8.21	136	348015	20.00	ng	-0.01
38) Acenaphthene-d10	9.97	164	234443	20.00	ng	-0.01
63) Phenanthrene-d10	11.45	188	490287	20.00	ng	-0.01
75) Chrysene-d12	14.09	240	397162	20.00	ng	-0.02
86) Perylene-d12	15.57	264	377988	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.51	112	8427	1.33	ng	-0.01
7) Phenol-d6	6.55	99	6519	0.82	ng	-0.01
23) Nitrobenzene-d5	7.49	82	14212	2.72	ng	-0.01
41) 2,4,6-Tribromophenol	10.75	330	9928	4.18	ng	-0.01
44) 2-Fluorobiphenyl	9.28	172	60631	2.69	ng	-0.02
78) Terphenyl-d14	13.04	244	63147	2.27	ng	-0.01
Target Compounds						
31) Naphthalene	8.23	128	900999	58.43	ng	Qvalue 97
37) 2-Methylnaphthalene	8.91	142	87801	8.34	ng	# 89
45) 1,1'-Biphenyl	9.38	154	38211	2.11	ng	96
48) Acenaphthylene	9.83	152	140640	6.19	ng	98
70) Phenanthrene	11.48	178	85771	2.05	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF092915\
Data File : BF081903.D
Acq On : 30 Sep 2015 8:02
Operator : UM/IZ
Sample : G3754-02DL2 25X
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
MW-2BDL2

Quant Time: Sep 30 10:55:10 2015
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF092815.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Sep 28 18:10:42 2015
Response via : Initial Calibration

