


Data Path : Z:\HPCHEM1\BNA_F\DATA\BF092915\
 Data File : BF081875.D
 Acq On : 29 Sep 2015 12:49
 Operator : UM/IZ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF092815.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Sep 28 18:10:42 2015



AutoFind: Scans 836, 837, 838; Background Corrected with Scan 833

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	35.0	11367	PASS
68	69	0.00	2	1.4	158	PASS
69	198	0.00	100	33.7	10944	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	45.2	14690	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	32515	PASS
199	198	5	9	6.4	2095	PASS
275	198	10	60	26.5	8613	PASS
365	198	1	100	3.3	1058	PASS
441	442	0.01	24	16.3	4914	PASS
442	198	50	100	92.5	30091	PASS
443	442	15	24	19.2	5791	PASS