


Data Path : Z:\HPCHEM1\BNA F\Data\BF092917\
 Data File : BF098936.D
 Acq On : 29 Sep 2017 8:49
 Operator : SJ/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF092317.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Sat Sep 23 13:50:58 2017



AutoFind: Scans 1631, 1632, 1633; Background Corrected with Scan 1626

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	38.5	30673	PASS
68	69	0.00	2	1.6	450	PASS
69	198	0.00	100	34.4	27346	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	51.9	41275	PASS
197	198	0.00	2	1.1	897	PASS
198	198	100	100	100.0	79592	PASS
199	198	5	9	6.7	5344	PASS
275	198	10	60	22.1	17603	PASS
365	198	1	100	2.6	2052	PASS
441	442	0.01	24	15.9	9739	PASS
442	198	50	100	76.8	61154	PASS
443	442	15	24	19.0	11593	PASS