

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF092922\
 Data File : BF130462.D
 Acq On : 29 Sep 2022 11:30
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 30 03:06:00 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 28 02:59:31 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	104	0.00
2	1,4-Dioxane	40.000	34.298	14.3	89	-0.02
3	Pyridine	40.000	35.350	11.6	95	-0.02
4	n-Nitrosodimethylamine	40.000	31.638	20.9	81	-0.02
5 S	2-Fluorophenol	80.000	79.458	0.7	105	0.00
6	Aniline	40.000	38.704	3.2	102	0.00
7 S	Phenol-d6	80.000	79.564	0.5	105	0.00
8	2-Chlorophenol	40.000	41.119	-2.8	108	0.00
9	Benzaldehyde	40.000	35.963	10.1	97	0.00
10 C	Phenol	40.000	38.880	2.8	101	0.00
11	bis(2-Chloroethyl)ether	40.000	38.787	3.0	102	0.00
12	1,3-Dichlorobenzene	40.000	40.836	-2.1	108	0.00
13 C	1,4-Dichlorobenzene	40.000	41.406	-3.5	109	0.00
14	1,2-Dichlorobenzene	40.000	40.877	-2.2	107	0.00
15	Benzyl Alcohol	40.000	32.089	19.8	81	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	32.779	18.1	86	0.00
17	2-Methylphenol	40.000	40.952	-2.4	106	0.00
18	Hexachloroethane	40.000	39.176	2.1	102	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	36.777	8.1	95	0.00
20	3+4-Methylphenols	40.000	40.471	-1.2	105	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	104	0.00
22	Acetophenone	40.000	39.220	2.0	102	0.00
23 S	Nitrobenzene-d5	80.000	76.412	4.5	98	0.00
24	Nitrobenzene	40.000	37.778	5.6	98	0.00
25	Isophorone	40.000	38.917	2.7	101	0.00
26 C	2-Nitrophenol	40.000	46.294	-15.7	114	0.00
27	2,4-Dimethylphenol	40.000	42.086	-5.2	108	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.958	0.1	105	0.00
29 C	2,4-Dichlorophenol	40.000	44.240	-10.6	113	0.00
30	1,2,4-Trichlorobenzene	40.000	43.347	-8.4	113	0.00
31	Naphthalene	40.000	41.331	-3.3	108	0.00
32	Benzoic acid	40.000	35.134	12.2	85	0.00
33	4-Chloroaniline	40.000	43.397	-8.5	112	0.00
34 C	Hexachlorobutadiene	40.000	43.351	-8.4	113	0.00
35	Caprolactam	40.000	43.479	-8.7	108	0.00
36 C	4-Chloro-3-methylphenol	40.000	41.737	-4.3	107	0.00
37	2-Methylnaphthalene	40.000	42.458	-6.1	111	0.00
38	1-Methylnaphthalene	40.000	42.772	-6.9	112	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	109	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	42.550	-6.4	119	0.00
41 P	Hexachlorocyclopentadiene	40.000	38.584	3.5	101	0.00
42 S	2,4,6-Tribromophenol	80.000	90.810	-13.5	122	0.00
43 C	2,4,6-Trichlorophenol	40.000	41.711	-4.3	112	0.00
44	2,4,5-Trichlorophenol	40.000	43.255	-8.1	119	0.00
45 S	2-Fluorobiphenyl	80.000	80.291	-0.4	112	0.00
46	1,1'-Biphenyl	40.000	40.334	-0.8	112	0.00
47	2-Chloronaphthalene	40.000	40.806	-2.0	113	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF092922\
 Data File : BF130462.D
 Acq On : 29 Sep 2022 11:30
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 30 03:06:00 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 28 02:59:31 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	36.733	8.2	97	0.00
49	Acenaphthylene	40.000	40.591	-1.5	113	0.00
50	Dimethylphthalate	40.000	41.380	-3.5	114	0.00
51	2,6-Dinitrotoluene	40.000	44.560	-11.4	122	0.00
52 C	Acenaphthene	40.000	40.752	-1.9	113	0.00
53	3-Nitroaniline	40.000	42.471	-6.2	114	0.00
54 P	2,4-Dinitrophenol	40.000	38.420	3.9	107	0.00
55	Dibenzofuran	40.000	41.012	-2.5	114	0.00
56 P	4-Nitrophenol	40.000	35.830	10.4	92	0.00
57	2,4-Dinitrotoluene	40.000	44.463	-11.2	117	0.00
58	Fluorene	40.000	40.975	-2.4	112	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	43.946	-9.9	121	0.00
60	Diethylphthalate	40.000	40.669	-1.7	111	0.00
61	4-Chlorophenyl-phenylether	40.000	42.900	-7.2	118	0.00
62	4-Nitroaniline	40.000	41.374	-3.4	110	0.00
63	Azobenzene	40.000	36.340	9.1	99	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	109	0.00
65	4,6-Dinitro-2-methylphenol	40.000	46.461	-16.2	123	0.00
66 c	n-Nitrosodiphenylamine	40.000	41.243	-3.1	115	0.00
67	4-Bromophenyl-phenylether	40.000	44.634	-11.6	125	0.00
68	Hexachlorobenzene	40.000	44.379	-10.9	124	0.00
69	Atrazine	40.000	40.906	-2.3	113	0.00
70 C	Pentachlorophenol	40.000	33.066	17.3	86	0.00
71	Phenanthrene	40.000	40.687	-1.7	114	0.00
72	Anthracene	40.000	40.901	-2.3	113	0.00
73	Carbazole	40.000	40.251	-0.6	111	0.00
74	Di-n-butylphthalate	40.000	41.421	-3.6	112	0.00
75 C	Fluoranthene	40.000	41.263	-3.2	113	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	105	0.00
77	Benzidine	40.000	35.873	10.3	89	0.00
78	Pyrene	40.000	42.415	-6.0	111	0.00
79 S	Terphenyl-d14	80.000	88.760	-11.0	116	0.00
80	Butylbenzylphthalate	40.000	45.199	-13.0	115	0.00
81	Benzo(a)anthracene	40.000	41.239	-3.1	110	0.00
82	3,3'-Dichlorobenzidine	40.000	45.405	-13.5	118	0.00
83	Chrysene	40.000	41.267	-3.2	111	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	47.053	-17.6	121	0.00
85 c	Di-n-octyl phthalate	40.000	43.325	-8.3	115	0.00
86 I	Perylene-d12	20.000	20.000	0.0	112	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	43.449	-8.6	117	0.00
88	Benzo(b)fluoranthene	40.000	40.740	-1.9	115	0.00
89	Benzo(k)fluoranthene	40.000	41.566	-3.9	118	0.00
90 C	Benzo(a)pyrene	40.000	43.068	-7.7	120	0.00
91	Dibenzo(a,h)anthracene	40.000	43.731	-9.3	119	0.00
92	Benzo(g,h,i)perylene	40.000	42.779	-6.9	116	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF092922\
Data File : BF130462.D
Acq On : 29 Sep 2022 11:30
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Sep 30 03:06:00 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091422.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 28 02:59:31 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0