

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF093015\
 Data File : BF081917.D
 Acq On : 30 Sep 2015 22:51
 Operator : UM/IZ
 Sample : G3868-05
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SB-01(WEST)

Manual Integrations
 APPROVED

MMDadoda
 10/1/2015 2:23:41 PM

Quant Time: Oct 01 04:49:44 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF092815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 01 04:39:45 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.90	152	106993m	20.00	ng	0.00
21) Naphthalene-d8	8.19	136	404956	20.00	ng	0.00
38) Acenaphthene-d10	9.94	164	184849	20.00	ng	-0.01
63) Phenanthrene-d10	11.43	188	344621	20.00	ng	-0.01
75) Chrysene-d12	14.07	240	269559	20.00	ng	-0.01
86) Perylene-d12	15.52	264	258618	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.50	112	488492	82.88	ng	0.00
7) Phenol-d6	6.53	99	637902	86.01	ng	-0.01
23) Nitrobenzene-d5	7.47	82	398619	65.62	ng	0.00
41) 2,4,6-Tribromophenol	10.74	330	147818	78.96	ng	0.00
44) 2-Fluorobiphenyl	9.27	172	817746	71.83	ng	0.00
78) Terphenyl-d14	13.02	244	690971	81.98	ng	0.00
Target Compounds						Qvalue
49) Dimethylphthalate	9.66	163	161849	12.68	ng	98
74) Fluoranthene	12.64	202	41380	2.24	ng	93
77) Pyrene	12.87	202	35328m	2.19	ng	
83) Bis(2-ethylhexyl)phthalate	14.05	149	17762	2.34	ng	# 89
84) Di-n-octyl phthalate	14.70	149	265147	21.43	ng	# 66

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF093015\
Data File : BF081917.D
Acq On : 30 Sep 2015 22:51
Operator : UM/IZ
Sample : G3868-05
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SB-01(WEST)

Manual Integrations
APPROVED

MMDadoda
10/1/2015 2:23:41 PM

Quant Time: Oct 01 04:49:44 2015
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF092815.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 01 04:39:45 2015
Response via : Initial Calibration

