

Data Path : Z:\HPCHEM1\BNA F\DATA\BF093017\
 Data File : BF098963.D
 Acq On : 30 Sep 2017 15:27
 Operator : SJ/JU
 Sample : I5487-07 5X
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 CL-02-092717-C

Manual Integrations
 APPROVED

Sohil
 10/2/2017 5:44:17 PM

Quant Time: Oct 01 00:33:11 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF092317.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Sep 23 13:50:58 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.92	152	137585	20.00	ng	0.00
21) Naphthalene-d8	8.19	136	561889	20.00	ng	-0.01
38) Acenaphthene-d10	9.95	164	243244	20.00	ng	-0.01
63) Phenanthrene-d10	11.44	188	376267	20.00	ng	0.00
75) Chrysene-d12	14.07	240	212824	20.00	ng	-0.01
86) Perylene-d12	15.57	264	234693	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.52	112	199139	23.04	ng	-0.02
7) Phenol-d6	6.53	99	233609	21.91	ng	-0.02
23) Nitrobenzene-d5	7.47	82	142946	16.31	ng	-0.02
41) 2,4,6-Tribromophenol	10.74	330	37278	18.73	ng	-0.01
44) 2-Fluorobiphenyl	9.27	172	279170	19.16	ng	-0.01
78) Terphenyl-d14	13.02	244	166421	17.78	ng	-0.02
Target Compounds						
49) Dimethylphthalate	9.66	163	39229	2.137	ng	99
85) Indeno(1,2,3-cd)pyrene	17.07	276	23732	2.418	ng	94
87) Benzo(b)fluoranthene	15.13	252	38418m	2.662	ng	
89) Benzo(a)pyrene	15.50	252	31056	2.345	ng	# 91
91) Benzo(g,h,i)perylene	17.52	276	24235	2.313	ng	# 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF093017\
Data File : BF098963.D
Acq On : 30 Sep 2017 15:27
Operator : SJ/JU
Sample : I5487-07 5X
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
CL-02-092717-C

Manual Integrations
APPROVED

Sohil
10/2/2017 5:44:17 PM

Quant Time: Oct 01 00:33:11 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF092317.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Sep 23 13:50:58 2017
Response via : Initial Calibration

