

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF093020\
 Data File : BF121744.D
 Acq On : 30 Sep 2020 18:02
 Operator : JU/CG
 Sample : PB132004BL
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB132004BL

Quant Time: Sep 30 18:37:35 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF091020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 29 11:12:14 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.77	152	147550	20.00	ng	-0.01
21) Naphthalene-d8	8.06	136	579835	20.00	ng	-0.01
39) Acenaphthene-d10	9.81	164	308190	20.00	ng	-0.02
64) Phenanthrene-d10	11.30	188	602177	20.00	ng	-0.01
76) Chrysene-d12	13.93	240	496406	20.00	ng	-0.01
86) Perylene-d12	15.37	264	393825	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.40	112	985109	99.22	ng	0.00
7) Phenol-d6	6.42	99	1243356	95.45	ng	-0.01
23) Nitrobenzene-d5	7.34	82	864228	80.03	ng	-0.02
42) 2,4,6-Tribromophenol	10.60	330	412421	107.67	ng	-0.01
45) 2-Fluorobiphenyl	9.13	172	1562880	76.80	ng	-0.02
79) Terphenyl-d14	12.89	244	1791043	73.09	ng	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF093020\
Data File : BF121744.D
Acq On : 30 Sep 2020 18:02
Operator : JU/CG
Sample : PB132004BL
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB132004BL

Quant Time: Sep 30 18:37:35 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF091020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 29 11:12:14 2020
Response via : Initial Calibration

