

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF093022\
 Data File : BF130480.D
 Acq On : 30 Sep 2022 12:19
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 01 01:34:08 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 28 02:59:31 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	116	0.00
2	1,4-Dioxane	40.000	35.389	11.5	103	0.02
3	Pyridine	40.000	38.315	4.2	115	0.02
4	n-Nitrosodimethylamine	40.000	32.662	18.3	94	0.03
5 S	2-Fluorophenol	80.000	79.972	0.0	117	0.01
6	Aniline	40.000	38.433	3.9	113	0.00
7 S	Phenol-d6	80.000	79.643	0.4	117	0.00
8	2-Chlorophenol	40.000	40.408	-1.0	119	0.01
9	Benzaldehyde	40.000	35.400	11.5	106	0.00
10 C	Phenol	40.000	38.621	3.4	112	0.00
11	bis(2-Chloroethyl)ether	40.000	38.824	2.9	114	0.01
12	1,3-Dichlorobenzene	40.000	40.417	-1.0	120	0.00
13 C	1,4-Dichlorobenzene	40.000	40.445	-1.1	119	0.00
14	1,2-Dichlorobenzene	40.000	40.081	-0.2	117	0.00
15	Benzyl Alcohol	40.000	35.086	12.3	99	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	32.776	18.1	96	0.00
17	2-Methylphenol	40.000	39.897	0.3	115	0.01
18	Hexachloroethane	40.000	38.649	3.4	112	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	36.124	9.7	104	0.00
20	3+4-Methylphenols	40.000	40.176	-0.4	117	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	113	0.00
22	Acetophenone	40.000	39.729	0.7	112	0.00
23 S	Nitrobenzene-d5	80.000	77.454	3.2	108	0.00
24	Nitrobenzene	40.000	38.226	4.4	107	0.00
25	Isophorone	40.000	38.845	2.9	109	0.00
26 C	2-Nitrophenol	40.000	47.238	-18.1	127	0.00
27	2,4-Dimethylphenol	40.000	41.378	-3.4	116	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.445	-1.1	115	0.00
29 C	2,4-Dichlorophenol	40.000	44.588	-11.5	124	0.00
30	1,2,4-Trichlorobenzene	40.000	43.094	-7.7	122	0.00
31	Naphthalene	40.000	40.751	-1.9	116	0.00
32	Benzoic acid	40.000	36.750	8.1	96	0.01
33	4-Chloroaniline	40.000	41.986	-5.0	118	0.00
34 C	Hexachlorobutadiene	40.000	42.562	-6.4	121	0.01
35	Caprolactam	40.000	44.558	-11.4	121	0.01
36 C	4-Chloro-3-methylphenol	40.000	41.773	-4.4	116	0.01
37	2-Methylnaphthalene	40.000	42.032	-5.1	119	0.00
38	1-Methylnaphthalene	40.000	42.351	-5.9	121	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	118	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	42.153	-5.4	127	0.00
41 P	Hexachlorocyclopentadiene	40.000	35.576	11.1	101	0.01
42 S	2,4,6-Tribromophenol	80.000	88.451	-10.6	128	0.00
43 C	2,4,6-Trichlorophenol	40.000	41.477	-3.7	120	0.00
44	2,4,5-Trichlorophenol	40.000	41.920	-4.8	124	0.00
45 S	2-Fluorobiphenyl	80.000	80.449	-0.6	121	0.00
46	1,1'-Biphenyl	40.000	40.207	-0.5	120	0.00
47	2-Chloronaphthalene	40.000	40.276	-0.7	120	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF093022\
 Data File : BF130480.D
 Acq On : 30 Sep 2022 12:19
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 01 01:34:08 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 28 02:59:31 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	36.851	7.9	105	0.00
49	Acenaphthylene	40.000	40.606	-1.5	121	0.00
50	Dimethylphthalate	40.000	41.273	-3.2	122	0.01
51	2,6-Dinitrotoluene	40.000	43.393	-8.5	128	0.00
52 C	Acenaphthene	40.000	40.563	-1.4	121	0.01
53	3-Nitroaniline	40.000	42.334	-5.8	122	0.00
54 P	2,4-Dinitrophenol	40.000	36.604	8.5	109	0.00
55	Dibenzofuran	40.000	40.677	-1.7	122	0.01
56 P	4-Nitrophenol	40.000	35.052	12.4	97	0.00
57	2,4-Dinitrotoluene	40.000	44.088	-10.2	125	0.00
58	Fluorene	40.000	40.263	-0.7	118	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	42.439	-6.1	126	0.00
60	Diethylphthalate	40.000	39.881	0.3	117	0.00
61	4-Chlorophenyl-phenylether	40.000	41.987	-5.0	125	0.00
62	4-Nitroaniline	40.000	41.089	-2.7	118	0.00
63	Azobenzene	40.000	36.403	9.0	107	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	116	0.00
65	4,6-Dinitro-2-methylphenol	40.000	47.091	-17.7	132	0.00
66 c	n-Nitrosodiphenylamine	40.000	41.869	-4.7	124	0.01
67	4-Bromophenyl-phenylether	40.000	44.716	-11.8	132	0.00
68	Hexachlorobenzene	40.000	43.874	-9.7	129	0.00
69	Atrazine	40.000	41.552	-3.9	121	0.00
70 C	Pentachlorophenol	40.000	35.042	12.4	96	0.00
71	Phenanthrene	40.000	40.802	-2.0	121	0.01
72	Anthracene	40.000	41.353	-3.4	121	0.00
73	Carbazole	40.000	41.126	-2.8	120	0.00
74	Di-n-butylphthalate	40.000	41.665	-4.2	119	0.00
75 C	Fluoranthene	40.000	41.232	-3.1	119	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	110	0.01
77	Benzydine	40.000	44.516	-11.3	116	0.00
78	Pyrene	40.000	43.812	-9.5	121	0.00
79 S	Terphenyl-d14	80.000	88.036	-10.0	121	0.00
80	Butylbenzylphthalate	40.000	43.987	-10.0	117	0.00
81	Benzo(a)anthracene	40.000	41.735	-4.3	117	0.00
82	3,3'-Dichlorobenzidine	40.000	45.884	-14.7	125	0.00
83	Chrysene	40.000	41.712	-4.3	118	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	43.122	-7.8	117	0.00
85 c	Di-n-octyl phthalate	40.000	43.149	-7.9	121	0.00
86 I	Perylene-d12	20.000	20.000	0.0	122	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	43.578	-8.9	128	0.01
88	Benzo(b)fluoranthene	40.000	42.896	-7.2	131	0.00
89	Benzo(k)fluoranthene	40.000	39.005	2.5	120	0.00
90 C	Benzo(a)pyrene	40.000	42.937	-7.3	130	0.01
91	Dibenzo(a,h)anthracene	40.000	43.591	-9.0	129	0.00
92	Benzo(g,h,i)perylene	40.000	41.991	-5.0	123	0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF093022\
Data File : BF130480.D
Acq On : 30 Sep 2022 12:19
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Oct 01 01:34:08 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091422.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 28 02:59:31 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0