

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF093024\
 Data File : BF139570.D
 Acq On : 30 Sep 2024 11:03
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 30 13:49:54 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Sep 14 02:58:23 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	185	-0.01
2	1,4-Dioxane	40.000	39.694	0.8	183	-0.04
3	Pyridine	40.000	43.650	-9.1	199	-0.04
4	n-Nitrosodimethylamine	40.000	34.209	14.5	158	-0.04
5 S	2-Fluorophenol	80.000	74.572	6.8	174	-0.01
6	Aniline	40.000	42.450	-6.1	199	-0.04
7 S	Phenol-d6	80.000	75.971	5.0	177	0.00
8	2-Chlorophenol	40.000	38.727	3.2	182	0.00
9	Benzaldehyde	40.000	49.383	-23.5	255	0.00
10 C	Phenol	40.000	39.376	1.6	185	0.00
11	bis(2-Chloroethyl)ether	40.000	38.942	2.6	198	-0.01
12	1,3-Dichlorobenzene	40.000	37.914	5.2	178	-0.01
13 C	1,4-Dichlorobenzene	40.000	38.716	3.2	181	-0.01
14	1,2-Dichlorobenzene	40.000	37.531	6.2	177	-0.01
15	Benzyl Alcohol	40.000	36.196	9.5	168	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	45.000	-12.5	214	0.00
17	2-Methylphenol	40.000	40.311	-0.8	189	0.00
18	Hexachloroethane	40.000	36.715	8.2	173	-0.01
19 P	n-Nitroso-di-n-propylamine	40.000	39.113	2.2	184	-0.01
20	3+4-Methylphenols	40.000	36.959	7.6	173	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	196	-0.01
22	Acetophenone	40.000	36.484	8.8	174	0.00
23 S	Nitrobenzene-d5	80.000	74.584	6.8	177	-0.01
24	Nitrobenzene	40.000	37.049	7.4	176	-0.01
25	Isophorone	40.000	40.138	-0.3	193	-0.01
26 C	2-Nitrophenol	40.000	41.437	-3.6	191	-0.01
27	2,4-Dimethylphenol	40.000	39.588	1.0	185	0.00
28	bis(2-Chloroethoxy)methane	40.000	41.552	-3.9	202	-0.01
29 C	2,4-Dichlorophenol	40.000	39.628	0.9	188	0.00
30	1,2,4-Trichlorobenzene	40.000	38.210	4.5	183	-0.01
31	Naphthalene	40.000	38.814	3.0	186	-0.01
32	Benzoic acid	40.000	40.837	-2.1	192	0.00
33	4-Chloroaniline	40.000	42.306	-5.8	196	-0.01
34 C	Hexachlorobutadiene	40.000	35.947	10.1	172	-0.01
35	Caprolactam	40.000	42.508	-6.3	196	0.00
36 C	4-Chloro-3-methylphenol	40.000	38.397	4.0	181	0.00
37	2-Methylnaphthalene	40.000	39.069	2.3	187	-0.02
38	1-Methylnaphthalene	40.000	38.862	2.8	187	-0.02
39 I	Acenaphthene-d10	20.000	20.000	0.0	204	-0.01
40	1,2,4,5-Tetrachlorobenzene	40.000	37.427	6.4	183	-0.01
41 P	Hexachlorocyclopentadiene	40.000	32.894	17.8	152	-0.01
42 S	2,4,6-Tribromophenol	80.000	69.921	12.6	166	0.00
43 C	2,4,6-Trichlorophenol	40.000	37.772	5.6	177	0.00
44	2,4,5-Trichlorophenol	40.000	38.504	3.7	184	0.00
45 S	2-Fluorobiphenyl	80.000	72.083	9.9	176	-0.01
46	1,1'-Biphenyl	40.000	38.271	4.3	187	-0.01
47	2-Chloronaphthalene	40.000	38.240	4.4	185	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF093024\
 Data File : BF139570.D
 Acq On : 30 Sep 2024 11:03
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 30 13:49:54 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Sep 14 02:58:23 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	38.126	4.7	183	-0.01
49	Acenaphthylene	40.000	38.727	3.2	189	-0.01
50	Dimethylphthalate	40.000	38.743	3.1	188	-0.01
51	2,6-Dinitrotoluene	40.000	40.755	-1.9	194	-0.01
52 C	Acenaphthene	40.000	37.754	5.6	181	-0.01
53	3-Nitroaniline	40.000	39.680	0.8	185	0.00
54 P	2,4-Dinitrophenol	40.000	36.663	8.3	170	0.00
55	Dibenzofuran	40.000	37.974	5.1	184	-0.01
56 P	4-Nitrophenol	40.000	37.237	6.9	173	0.00
57	2,4-Dinitrotoluene	40.000	39.090	2.3	181	0.00
58	Fluorene	40.000	36.727	8.2	175	-0.01
59	2,3,4,6-Tetrachlorophenol	40.000	36.035	9.9	172	-0.01
60	Diethylphthalate	40.000	38.037	4.9	183	-0.01
61	4-Chlorophenyl-phenylether	40.000	37.115	7.2	180	-0.01
62	4-Nitroaniline	40.000	36.327	9.2	169	-0.01
63	Azobenzene	40.000	38.895	2.8	186	-0.01
64 I	Phenanthrene-d10	20.000	20.000	0.0	192	-0.01
65	4,6-Dinitro-2-methylphenol	40.000	42.962	-7.4	187	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.921	0.2	187	0.00
67	4-Bromophenyl-phenylether	40.000	40.225	-0.6	185	-0.02
68	Hexachlorobenzene	40.000	37.652	5.9	173	-0.01
69	Atrazine	40.000	44.258	-10.6	244	0.00
70 C	Pentachlorophenol	40.000	37.182	7.0	166	0.00
71	Phenanthrene	40.000	38.486	3.8	178	-0.01
72	Anthracene	40.000	38.730	3.2	178	-0.01
73	Carbazole	40.000	38.826	2.9	178	-0.01
74	Di-n-butylphthalate	40.000	41.690	-4.2	187	-0.01
75 C	Fluoranthene	40.000	37.118	7.2	169	-0.01
76 I	Chrysene-d12	20.000	20.000	0.0	169	-0.01
77	Benzidine	40.000	37.459	6.4	118	-0.01
78	Pyrene	40.000	44.879	-12.2	169	0.00
79 S	Terphenyl-d14	80.000	83.947	-4.9	160	-0.01
80	Butylbenzylphthalate	40.000	45.357	-13.4	178	-0.01
81	Benzo(a)anthracene	40.000	39.631	0.9	163	-0.01
82	3,3'-Dichlorobenzidine	40.000	40.255	-0.6	176	-0.01
83	Chrysene	40.000	39.646	0.9	165	-0.01
84	Bis(2-ethylhexyl)phthalate	40.000	46.486	-16.2	191	-0.01
85 c	Di-n-octyl phthalate	40.000	45.070	-12.7	184	-0.01
86 I	Perylene-d12	20.000	20.000	0.0	174	-0.02
87	Indeno(1,2,3-cd)pyrene	40.000	42.902	-7.3	162	-0.02
88	Benzo(b)fluoranthene	40.000	37.462	6.3	164	-0.01
89	Benzo(k)fluoranthene	40.000	38.074	4.8	158	-0.01
90 C	Benzo(a)pyrene	40.000	39.275	1.8	162	-0.01
91	Dibenzo(a,h)anthracene	40.000	42.845	-7.1	166	-0.02
92	Benzo(g,h,i)perylene	40.000	44.781	-12.0	171	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF093024\
Data File : BF139570.D
Acq On : 30 Sep 2024 11:03
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Sep 30 13:49:54 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091324.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Sep 14 02:58:23 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0