

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF100115\
 Data File : BF081945.D
 Acq On : 1 Oct 2015 18:10
 Operator : UM/IZ
 Sample : G3707-05
 Misc : LOD 2 PPM
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 LOD-WATER2015-01

Manual Integrations
 APPROVED

apatel
 10/2/2015 9:02:16 AM

Quant Time: Oct 02 05:03:17 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF092815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Sep 28 18:10:42 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.89	152	119484	20.00	ng	-0.05
21) Naphthalene-d8	8.18	136	486886	20.00	ng	-0.03
38) Acenaphthene-d10	9.93	164	237483	20.00	ng	-0.05
63) Phenanthrene-d10	11.42	188	465345	20.00	ng	-0.05
75) Chrysene-d12	14.06	240	477339	20.00	ng	-0.06
86) Perylene-d12	15.50	264	376251	20.00	ng	-0.08
System Monitoring Compounds						
5) 2-Fluorophenol	5.50	112	643808	97.81	ng	-0.02
7) Phenol-d6	6.53	99	915231	110.50	ng	-0.03
23) Nitrobenzene-d5	7.46	82	547819	75.00	ng	-0.03
41) 2,4,6-Tribromophenol	10.72	330	239427	99.55	ng	-0.05
44) 2-Fluorobiphenyl	9.26	172	1088084	74.92	ng	-0.05
78) Terphenyl-d14	13.01	244	1143939	74.90	ng	-0.05

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF100115\
Data File : BF081945.D
Acq On : 1 Oct 2015 18:10
Operator : UM/IZ
Sample : G3707-05
Misc : LOD 2 PPM
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
LOD-WATER2015-01

Manual Integrations
APPROVED

apatel
10/2/2015 9:02:16 AM

Quant Time: Oct 02 05:03:17 2015
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF092815.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Sep 28 18:10:42 2015
Response via : Initial Calibration

