

Data Path : Z:\HPCHEM1\BNA\_F\DATA\BF100115\  
 Data File : BF081978.D  
 Acq On : 2 Oct 2015 10:52  
 Operator : UM/IZ  
 Sample : G3875-08  
 Misc :  
 ALS Vial : 42 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 A8-COMP

Quant Time: Oct 02 12:51:05 2015  
 Quant Method : Z:\HPCHEM1\BNA\_F\METHODS\8270-BF092815.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Mon Sep 28 18:10:42 2015  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc   | Units | Dev(Min)       |
|-----------------------------|-------|------|----------|--------|-------|----------------|
| 1) 1,4-Dichlorobenzene-d4   | 6.89  | 152  | 95712    | 20.00  | ng    | -0.05          |
| 21) Naphthalene-d8          | 8.18  | 136  | 384913   | 20.00  | ng    | -0.03          |
| 38) Acenaphthene-d10        | 9.93  | 164  | 193302   | 20.00  | ng    | -0.05          |
| 63) Phenanthrene-d10        | 11.42 | 188  | 378311   | 20.00  | ng    | -0.05          |
| 75) Chrysene-d12            | 14.06 | 240  | 339143   | 20.00  | ng    | -0.06          |
| 86) Perylene-d12            | 15.50 | 264  | 240475   | 20.00  | ng    | -0.08          |
| System Monitoring Compounds |       |      |          |        |       |                |
| 5) 2-Fluorophenol           | 5.50  | 112  | 676042   | 128.22 | ng    | -0.02          |
| 7) Phenol-d6                | 6.53  | 99   | 915164   | 137.94 | ng    | -0.03          |
| 23) Nitrobenzene-d5         | 7.46  | 82   | 541436   | 93.77  | ng    | -0.03          |
| 41) 2,4,6-Tribromophenol    | 10.72 | 330  | 238230   | 121.69 | ng    | -0.05          |
| 44) 2-Fluorobiphenyl        | 9.26  | 172  | 1001390  | 87.10  | ng    | -0.05          |
| 78) Terphenyl-d14           | 13.01 | 244  | 1001847  | 100.35 | ng    | -0.05          |
| Target Compounds            |       |      |          |        |       |                |
| 49) Dimethylphthalate       | 9.65  | 163  | 237340   | 17.79  | ng    | Qvalue<br># 97 |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA\_F\DATA\BF100115\  
Data File : BF081978.D  
Acq On : 2 Oct 2015 10:52  
Operator : UM/IZ  
Sample : G3875-08  
Misc :  
ALS Vial : 42 Sample Multiplier: 1

Instrument :  
BNA\_F  
ClientSampleId :  
A8-COMP

Quant Time: Oct 02 12:51:05 2015  
Quant Method : Z:\HPCHEM1\BNA\_F\METHODS\8270-BF092815.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Mon Sep 28 18:10:42 2015  
Response via : Initial Calibration

