


Data Path : Z:\HPCHEM1\BNA_F\DATA\BF100115\
 Data File : BF081950.D
 Acq On : 1 Oct 2015 20:42
 Operator : UM/IZ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF092815.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Sep 28 18:10:42 2015



AutoFind: Scans 833, 834, 835; Background Corrected with Scan 830

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	33.5	17715	PASS
68	69	0.00	2	1.2	209	PASS
69	198	0.00	100	31.9	16841	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	44.0	23243	PASS
197	198	0.00	2	0.4	216	PASS
198	198	100	100	100.0	52807	PASS
199	198	5	9	7.1	3733	PASS
275	198	10	60	24.8	13102	PASS
365	198	1	100	2.8	1499	PASS
441	442	0.01	24	15.6	7774	PASS
442	198	50	100	94.2	49753	PASS
443	442	15	24	19.1	9492	PASS