

Data Path : Z:\HPCHEM1\BNA F\DATA\BF100215\
 Data File : BF082031.D
 Acq On : 3 Oct 2015 19:56
 Operator : UM/IZ
 Sample : G3867-06
 Misc :
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SB-06-6.5-7.0-092915

Quant Time: Oct 05 04:22:17 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF092815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Sep 28 18:10:42 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.88	152	76832	20.00	ng	-0.06
21) Naphthalene-d8	8.17	136	327141	20.00	ng	-0.05
38) Acenaphthene-d10	9.93	164	167554	20.00	ng	-0.05
63) Phenanthrene-d10	11.42	188	325106	20.00	ng	-0.05
75) Chrysene-d12	14.07	240	300742	20.00	ng	-0.05
86) Perylene-d12	15.53	264	226970	20.00	ng	-0.05
System Monitoring Compounds						
5) 2-Fluorophenol	5.47	112	600853	141.96	ng	-0.04
7) Phenol-d6	6.51	99	803510	150.87	ng	-0.05
23) Nitrobenzene-d5	7.45	82	487972	99.43	ng	-0.05
41) 2,4,6-Tribromophenol	10.73	330	265708	156.58	ng	-0.03
44) 2-Fluorobiphenyl	9.25	172	727384	70.23	ng	-0.06
78) Terphenyl-d14	13.02	244	847093	93.55	ng	-0.03
Target Compounds						
37) 2-Methylnaphthalene	8.98	142	36399	3.68	ng	# 87
49) Dimethylphthalate	9.65	163	268313	23.20	ng	# 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF100215\
Data File : BF082031.D
Acq On : 3 Oct 2015 19:56
Operator : UM/IZ
Sample : G3867-06
Misc :
ALS Vial : 51 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SB-06-6.5-7.0-092915

Quant Time: Oct 05 04:22:17 2015
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF092815.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Sep 28 18:10:42 2015
Response via : Initial Calibration

