

Data Path : Z:\HPCHEM1\BNA F\DATA\BF100215\
 Data File : BF082036.D
 Acq On : 3 Oct 2015 22:25
 Operator : UM/IZ
 Sample : G3846-16
 Misc :
 ALS Vial : 56 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 016-TB-13(4-8)

Manual Integrations
 APPROVED

MMDADODA
 10/5/2015 6:27:20 PM

Quant Time: Oct 05 17:46:46 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF092815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Sep 28 18:10:42 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.90	152	65521	20.00	ng	-0.03
21) Naphthalene-d8	8.22	136	141173	20.00	ng	0.00
38) Acenaphthene-d10	10.00	164	28564	20.00	ng	0.02
63) Phenanthrene-d10	11.58	188	100914m	20.00	ng	0.11
75) Chrysene-d12	14.20	240	215959	20.00	ng	0.08
86) Perylene-d12	15.60	264	204001	20.00	ng	0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.47	112	532387	147.50	ng	-0.04
7) Phenol-d6	6.51	99	638968	140.69	ng	-0.05
23) Nitrobenzene-d5	7.47	82	281585	132.96	ng	-0.02
41) 2,4,6-Tribromophenol	10.80	330	35888	124.06	ng	0.03
44) 2-Fluorobiphenyl	9.33	172	412192m	-20.00	ng	0.02
78) Terphenyl-d14	13.18	244	542253	79.71	ng	0.13
Target Compounds						Qvalue
31) Naphthalene	8.31	128	3088141m	493.67	ng	
37) 2-Methylnaphthalene	9.04	142	7676393m	1797.77	ng	
45) 1,1'-Biphenyl	9.42	154	565199	256.48	ng	91
70) Phenanthrene	11.50	178	487864	102.59	ng	# 89
71) Anthracene	11.61	178	2555828m	498.29	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF100215\
Data File : BF082036.D
Acq On : 3 Oct 2015 22:25
Operator : UM/IZ
Sample : G3846-16
Misc :
ALS Vial : 56 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
016-TB-13(4-8)

Manual Integrations
APPROVED

MMDADODA
10/5/2015 6:27:20 PM

Quant Time: Oct 05 17:46:46 2015
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF092815.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Sep 28 18:10:42 2015
Response via : Initial Calibration

