

Data Path : Z:\HPCHEM1\BNA F\DATA\BF100217\
 Data File : BF099020.D
 Acq On : 3 Oct 2017 00:42
 Operator : SJ/JU
 Sample : I5531-01 5X
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 G-62-0-2.5

Quant Time: Oct 03 04:54:43 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF092317.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Sep 23 13:50:58 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.90	152	120266	20.00	ng	-0.02
21) Naphthalene-d8	8.19	136	458735	20.00	ng	-0.02
38) Acenaphthene-d10	9.95	164	174273	20.00	ng	-0.02
63) Phenanthrene-d10	11.43	188	268566	20.00	ng	-0.01
75) Chrysene-d12	14.07	240	251088	20.00	ng	-0.01
86) Perylene-d12	15.57	264	209262	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.52	112	73749	9.76	ng	-0.02
7) Phenol-d6	6.52	99	174692	18.74	ng	-0.02
23) Nitrobenzene-d5	7.46	82	110455	15.44	ng	-0.02
41) 2,4,6-Tribromophenol	10.73	330	1494	1.05	ng	-0.02
44) 2-Fluorobiphenyl	9.26	172	209401	20.06	ng	-0.02
78) Terphenyl-d14	13.01	244	149207	13.51	ng	-0.02
Target Compounds						
49) Dimethylphthalate	9.65	163	31071	2.362	ng	Qvalue # 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF100217\
Data File : BF099020.D
Acq On : 3 Oct 2017 00:42
Operator : SJ/JU
Sample : I5531-01 5X
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
G-62-0-2.5

Quant Time: Oct 03 04:54:43 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF092317.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Sep 23 13:50:58 2017
Response via : Initial Calibration

