

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF100218\
 Data File : BF109538.D
 Acq On : 3 Oct 2018 8:26
 Operator : JU/SJ
 Sample : J5187-17
 Misc :
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 092818-TIPC-F-D-B

Quant Time: Oct 03 09:29:48 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF092218.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Sep 22 13:32:18 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.70	152	98573	20.00	ng	0.00
21) Naphthalene-d8	7.98	136	377267	20.00	ng	-0.01
38) Acenaphthene-d10	9.73	164	179596	20.00	ng	-0.01
63) Phenanthrene-d10	11.21	188	329972	20.00	ng	0.00
75) Chrysene-d12	13.83	240	250706	20.00	ng	-0.01
86) Perylene-d12	15.23	264	244581	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.31	112	241176	40.30	ng	0.00
7) Phenol-d6	6.34	99	199813	26.16	ng	-0.02
23) Nitrobenzene-d5	7.27	82	678522	106.17	ng	-0.01
41) 2,4,6-Tribromophenol	10.52	330	185296	104.66	ng	-0.02
44) 2-Fluorobiphenyl	9.06	172	1231097	103.38	ng	-0.01
78) Terphenyl-d14	12.79	244	1123917	110.02	ng	-0.01
Target Compounds						Qvalue
49) Dimethylphthalate	9.45	163	40186	3.076	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF100218\
Data File : BF109538.D
Acq On : 3 Oct 2018 8:26
Operator : JU/SJ
Sample : J5187-17
Misc :
ALS Vial : 45 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
092818-TIPC-F-D-B

Quant Time: Oct 03 09:29:48 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF092218.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Sep 22 13:32:18 2018
Response via : Initial Calibration

