

Data Path : Z:\HPCHEM1\BNA F\DATA\BF100317\
 Data File : BF099028.D
 Acq On : 3 Oct 2017 13:42
 Operator : SJ/JU
 Sample : I5275-05
 Misc : LOD 1 PPM
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 LOD-WATER-05-QT4-2017

Manual Integrations
 APPROVED

Sohil
 10/4/2017 7:21:17 PM

Quant Time: Oct 05 00:54:40 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF092317.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Sep 23 13:50:58 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.87	152	118355	20.00	ng	-0.05
21) Naphthalene-d8	8.16	136	501777	20.00	ng	-0.05
38) Acenaphthene-d10	9.91	164	231058	20.00	ng	-0.05
63) Phenanthrene-d10	11.40	188	418202	20.00	ng	-0.05
75) Chrysene-d12	14.03	240	307550	20.00	ng	-0.05
86) Perylene-d12	15.51	264	234557	20.00	ng	-0.06
System Monitoring Compounds						
5) 2-Fluorophenol	5.49	112	1062919	142.96	ng	-0.05
7) Phenol-d6	6.50	99	1279992	139.56	ng	-0.04
23) Nitrobenzene-d5	7.43	82	644064	82.32	ng	-0.05
41) 2,4,6-Tribromophenol	10.70	330	276578	146.28	ng	-0.05
44) 2-Fluorobiphenyl	9.23	172	1201648	86.82	ng	-0.05
78) Terphenyl-d14	12.98	244	1004755	74.30	ng	-0.05
Target Compounds						
15) Benzyl Alcohol	7.03	79	15024	2.233	ng	Qvalue # 47

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF100317\
Data File : BF099028.D
Acq On : 3 Oct 2017 13:42
Operator : SJ/JU
Sample : I5275-05
Misc : LOD 1 PPM
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
LOD-WATER-05-QT4-2017

Manual Integrations
APPROVED

Sohil
10/4/2017 7:21:17 PM

Quant Time: Oct 05 00:54:40 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF092317.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Sep 23 13:50:58 2017
Response via : Initial Calibration

