

Data Path : Z:\HPCHEM1\BNA F\DATA\BF100317\
 Data File : BF099057.D
 Acq On : 4 Oct 2017 3:43
 Operator : SJ/JU
 Sample : I5588-02
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SASP2P-ENV-114-4(0-5)

Manual Integrations
 APPROVED

Sohil
 10/4/2017 7:21:36 PM

Quant Time: Oct 04 08:32:42 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF092317.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Sep 23 13:50:58 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.87	152	121734	20.00	ng	-0.05
21) Naphthalene-d8	8.16	136	456707	20.00	ng	-0.05
38) Acenaphthene-d10	9.91	164	163249	20.00	ng	-0.05
63) Phenanthrene-d10	11.40	188	242359	20.00	ng	-0.05
75) Chrysene-d12	14.03	240	250764	20.00	ng	-0.05
86) Perylene-d12	15.51	264	243512	20.00	ng	-0.06
System Monitoring Compounds						
5) 2-Fluorophenol	5.49	112	714064	93.38	ng	-0.05
7) Phenol-d6	6.50	99	862448	91.42	ng	-0.05
23) Nitrobenzene-d5	7.43	82	512904	72.02	ng	-0.05
41) 2,4,6-Tribromophenol	10.70	330	107083	80.16	ng	-0.05
44) 2-Fluorobiphenyl	9.23	172	794739	81.27	ng	-0.05
78) Terphenyl-d14	12.98	244	564538	51.20	ng	-0.05
Target Compounds						
10) Phenol	6.51	94	21204	2.010	ng	96
49) Dimethylphthalate	9.62	163	136630	11.089	ng	99
74) Fluoranthene	12.61	202	41651	3.171	ng	97
77) Pyrene	12.84	202	42247	2.209	ng	98
87) Benzo(b)fluoranthene	15.07	252	34097m	2.277	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF100317\
Data File : BF099057.D
Acq On : 4 Oct 2017 3:43
Operator : SJ/JU
Sample : I5588-02
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SASP2P-ENV-114-4(0-5)

Manual Integrations
APPROVED

Sohil
10/4/2017 7:21:36 PM

Quant Time: Oct 04 08:32:42 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF092317.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Sep 23 13:50:58 2017
Response via : Initial Calibration

