

Data Path : Z:\HPCHEM1\BNA F\DATA\BF100417\
 Data File : BF099076.D
 Acq On : 4 Oct 2017 19:36
 Operator : SJ/JU
 Sample : I5526-04
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :

Quant Time: Oct 05 01:09:59 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF092317.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Sep 23 13:50:58 2017
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.86	152	132692	20.00	ng	-0.06
21) Naphthalene-d8	8.15	136	554134	20.00	ng	-0.05
38) Acenaphthene-d10	9.90	164	248962	20.00	ng	-0.06
63) Phenanthrene-d10	11.39	188	424293	20.00	ng	-0.05
75) Chrysene-d12	14.03	240	255343	20.00	ng	-0.06
86) Perylene-d12	15.50	264	227270	20.00	ng	-0.07
System Monitoring Compounds						
5) 2-Fluorophenol	5.48	112	563695	67.63	ng	-0.06
7) Phenol-d6	6.49	99	448798	43.65	ng	-0.05
23) Nitrobenzene-d5	7.43	82	779621	90.23	ng	-0.05
41) 2,4,6-Tribromophenol	10.70	330	236530	116.11	ng	-0.05
44) 2-Fluorobiphenyl	9.22	172	1353413	90.75	ng	-0.06
78) Terphenyl-d14	12.97	244	1047168	93.27	ng	-0.06
Target Compounds						Qvalue
2) 1,4-Dioxane	2.67	88	17967	4.51	ng	# 95
9) Benzaldehyde	6.63	77	18002	3.02	ng	83
10) Phenol	6.50	94	39096	3.40	ng	94
18) Hexachloroethane	7.40	117	8759	2.42	ng	# 1
19) n-Nitroso-di-n-propylamine	7.43	70	102141	15.56	ng	# 74
25) Isophorone	7.43	82	748518	41.90	ng	# 64
27) 2,4-Dimethylphenol	7.80	122	97918	11.00	ng	98
31) Naphthalene	8.17	128	1954596	75.10	ng	99
32) Benzoic acid	7.93	122	102169	13.97	ng	# 13
33) 4-Chloroaniline	8.17	127	265991	24.47	ng	# 34
37) 2-Methylnaphthalene	8.86	142	45443	2.69	ng	95
45) 1,1'-Biphenyl	9.32	154	110207	5.15	ng	98
49) Dimethylphthalate	9.62	163	38642	2.06	ng	97
51) Acenaphthene	9.94	154	339952	21.82	ng	99
53) 2,4-Dinitrophenol	9.95	184	224	5.93	ng	# 1
54) Dibenzofuran	10.11	168	322841	15.56	ng	97
57) Fluorene	10.45	166	222647	13.35	ng	98
66) 4-Bromophenyl-phenylether	10.70	248	15705	3.78	ng	# 1
70) Phenanthrene	11.42	178	313927	13.82	ng	98
71) Anthracene	11.46	178	49600	2.13	ng	98
72) Carbazole	11.62	167	312962	13.75	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF100417\
Data File : BF099076.D
Acq On : 4 Oct 2017 19:36
Operator : SJ/JU
Sample : I5526-04
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :

Quant Time: Oct 05 01:09:59 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF092317.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Sep 23 13:50:58 2017
Response via : Initial Calibration









#1

1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 6.86 min Scan# 812
Delta R.T. -0.06 min
Lab File: BF099076.D
Acq: 4 Oct 2017 19:36

Instrument :
BNA_F
ClientSampleId :




Tot Ion:152 Resp: 132692
Ion Ratio Lower Upper
152 100
150 158.2 124.6 186.8
115 63.4 50.1 75.1




#2

1,4-Dioxane
Concen: 4.51 ng
RT: 2.67 min Scan# 100
Delta R.T. -0.08 min
Lab File: BF099076.D
Acq: 4 Oct 2017 19:36

Tgt Ion: 88 Resp: 17967
Ion Ratio Lower Upper
88 100
58 76.2 62.6 93.8
43 37.6 24.6 37.0#





#5

2-Fluorophenol

Concen: 67.63 ng

RT: 5.48 min Scan# 577

Delta R.T. -0.06 min

Lab File: BF099076.D

Acq: 4 Oct 2017 19:36

Instrument :

BNA_F

ClientSampleId :


Tot Ion:112 Resp: 563695

Ion Ratio Lower Upper

112 100

64 57.4 47.9 71.9




63 29.3 23.7 35.5



Abundance Ion 112.00 (111.70 to 112.70): E

Ion 64.00 (63.70 to 64.70): BF0

Ion 63.00 (62.70 to 63.70): BF0



#7

Phenol-d6

Concen: 43.65 ng

RT: 6.49 min Scan# 749

Delta R.T. -0.05 min

Lab File: BF099076.D

Acq: 4 Oct 2017 19:36


Tgt Ion: 99 Resp: 448798

Ion Ratio Lower Upper

99 100

42 15.8 14.5 21.7



71 29.6 25.4 38.0




Abundance Ion 99.00 (98.70 to 99.70): BF0

Ion 42.00 (41.70 to 42.70): BF0

Ion 71.00 (70.70 to 71.70): BF0





#9

Benzaldehyde

Concen: 3.02 ng

RT: 6.63 min Scan# 773

Delta R.T. 0.16 min

Lab File: BF099076.D

Acq: 4 Oct 2017 19:36

Instrument :
BNA_F
ClientSampleId :


Tot Ion: 77 Resp: 18002

Ion Ratio Lower Upper




77 100

105 84.2 81.1 121.1

106 78.7 74.5 114.5



Abundance Ion 77.00 (76.70 to 77.70): BF0
Ion 105.00 (104.70 to 105.70): BF0
Ion 106.00 (105.70 to 106.70): BF0



#10

Phenol

Concen: 3.40 ng

RT: 6.50 min Scan# 751

Delta R.T. -0.05 min

Lab File: BF099076.D

Acq: 4 Oct 2017 19:36


Tgt Ion: 94 Resp: 39096

Ion Ratio Lower Upper



94 100


65 25.1 7.9 47.9

66 32.7 16.2 56.2



Abundance Ion 94.00 (93.70 to 94.70): BF0
Ion 65.00 (64.70 to 65.70): BF0
Ion 66.00 (65.70 to 66.70): BF0









#18
Hexachloroethane
Concen: 2.42 ng
RT: 7.40 min Scan# 903
Delta R.T. -0.04 min
Lab File: BF099076.D
Acq: 4 Oct 2017 19:36




Instrument :
BNA_F
ClientSampled :


Tot Ion:117 Resp: 8759
Ion Ratio Lower Upper
117 100
119 419.0 75.8 113.8#
201 0.0 75.7 113.5#



#19
n-Nitroso-di-n-propylamine
Concen: 15.56 ng
RT: 7.43 min Scan# 909
Delta R.T. 0.10 min
Lab File: BF099076.D
Acq: 4 Oct 2017 19:36

Tgt Ion: 70 Resp: 102141
Ion Ratio Lower Upper
70 100
42 43.0 47.5 71.3#
101 0.5 7.4 11.2#
130 3.2 16.6 25.0#





#21

Naphthalene-d8

Concen: 20.00 ng

RT: 8.15 min Scan# 1031

Delta R.T. -0.05 min

Lab File: BF099076.D

Acq: 4 Oct 2017 19:36

Instrument :

BNA_F

ClientSampleId :

Tot Ion:136 Resp: 554134


Ion Ratio Lower Upper

136 100

137 10.5 8.9 13.3

54 7.1 6.6 9.8

68 6.0 5.0 7.4






Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BF0

Ion 68.00 (67.70 to 68.70): BF0



#23

Nitrobenzene-d5

Concen: 90.23 ng

RT: 7.43 min Scan# 909

Delta R.T. -0.05 min

Lab File: BF099076.D

Acq: 4 Oct 2017 19:36


Tgt Ion: 82 Resp: 779621

Ion Ratio Lower Upper

82 100

128 49.8 36.1 54.1



54 46.9 41.4 62.2




Abundance Ion 82.00 (81.70 to 82.70): BF0

Ion 128.00 (127.70 to 128.70): E

Ion 54.00 (53.70 to 54.70): BF0





#25

Isophorone

Concen: 41.90 ng

RT: 7.43 min Scan# 909

Delta R.T. -0.31 min

Lab File: BF099076.D

Acq: 4 Oct 2017 19:36

Instrument :

BNA_F

ClientSampled :


Tot Ion: 82 Resp: 748518

Ion Ratio Lower Upper




82 100

95 0.0 5.2 7.8#

138 0.0 14.9 22.3#



Abundance Ion 82.00 (81.70 to 82.70): BF0
Ion 95.00 (94.70 to 95.70): BF0
Ion 138.00 (137.70 to 138.70): BF0



#27

2,4-Dimethylphenol

Concen: 11.00 ng

RT: 7.80 min Scan# 971

Delta R.T. -0.05 min

Lab File: BF099076.D

Acq: 4 Oct 2017 19:36


Tgt Ion: 122 Resp: 97918

Ion Ratio Lower Upper



122 100


107 111.8 91.2 136.8

121 56.6 47.2 70.8



Abundance Ion 122.00 (121.70 to 122.70): E
Ion 107.00 (106.70 to 107.70): E
Ion 121.00 (120.70 to 121.70): E





#31

Naphthalene

Concen: 75.10 ng

RT: 8.17 min Scan# 1035

Delta R.T. -0.05 min

Lab File: BF099076.D

Acq: 4 Oct 2017 19:36

Instrument :

BNA_F

ClientSampleId :


Tot Ion:128 Resp: 1954596

Ion Ratio Lower Upper

128 100

129 11.5 8.8 13.2




127 13.9 10.9 16.3



Abundance Ion 128.00 (127.70 to 128.70): E

Ion 129.00 (128.70 to 129.70): E

Ion 127.00 (126.70 to 127.70): E



#32

Benzoic acid

Concen: 13.97 ng

RT: 7.93 min Scan# 993

Delta R.T. -0.05 min

Lab File: BF099076.D

Acq: 4 Oct 2017 19:36


Tgt Ion:122 Resp: 102169

Ion Ratio Lower Upper

122 100

105 3.2 101.1 141.1#



77 32.3 70.7 110.7#




Abundance Ion 122.00 (121.70 to 122.70): E

Ion 105.00 (104.70 to 105.70): E

Ion 77.00 (76.70 to 77.70): BF0









#33
4-Chloroaniline
Concen: 24.47 ng
RT: 8.17 min Scan# 1035
Delta R.T. -0.10 min
Lab File: BF099076.D
Acq: 4 Oct 2017 19:36




Instrument :
BNA_F
ClientSampleId :


Tot Ion:127	Resp:	265991
Ion Ratio	Lower	Upper
127	100	
129	82.9	25.9 38.9#
65	0.0	25.0 37.4#
92	0.1	15.2 22.8#



#37
2-Methylnaphthalene
Concen: 2.69 ng
RT: 8.86 min Scan# 1151
Delta R.T. -0.06 min
Lab File: BF099076.D
Acq: 4 Oct 2017 19:36

Tgt Ion:142	Resp:	45443
Ion Ratio	Lower	Upper
142	100	
141	91.7	70.8 106.2
115	38.9	26.1 39.1








#38
 Acenaphthene-d10
 Concen: 20.00 ng
 RT: 9.90 min Scan# 1329
 Delta R.T. -0.06 min
 Lab File: BF099076.D
 Acq: 4 Oct 2017 19:36




Instrument :
 BNA_F
 ClientSampleId :


Tot Ion:164 Resp: 248962
 Ion Ratio Lower Upper
 164 100
 162 107.0 86.1 129.1
 160 47.6 38.3 57.5



#41
 2,4,6-Tribromophenol
 Concen: 116.11 ng
 RT: 10.70 min Scan# 1464
 Delta R.T. -0.05 min
 Lab File: BF099076.D
 Acq: 4 Oct 2017 19:36

Tgt Ion:330 Resp: 236530
 Ion Ratio Lower Upper
 330 100
 332 97.0 77.0 115.6
 141 39.3 29.9 44.9









#44
2-Fluorobiphenyl
Concen: 90.75 ng
RT: 9.22 min Scan# 1213
Delta R.T. -0.06 min
Lab File: BF099076.D
Acq: 4 Oct 2017 19:36




Instrument :
BNA_F
ClientSampleId :


Tot Ion:172 Resp: 1353413
Ion Ratio Lower Upper
172 100
171 36.9 29.7 44.5
170 24.2 19.6 29.4



#45
1,1'-Biphenyl
Concen: 5.15 ng
RT: 9.32 min Scan# 1230
Delta R.T. -0.06 min
Lab File: BF099076.D
Acq: 4 Oct 2017 19:36

Tgt Ion:154 Resp: 110207
Ion Ratio Lower Upper
154 100
153 39.9 21.3 61.3
76 13.5 0.0 34.0









#49
Dimethylphthalate
Concen: 2.06 ng
RT: 9.62 min Scan# 1280
Delta R.T. -0.06 min
Lab File: BF099076.D
Acq: 4 Oct 2017 19:36

Instrument :
BNA_F
ClientSampleId :

Tot Ion:163 Resp: 38642
Ion Ratio Lower Upper
163 100
194 3.3 2.7 4.1
164 11.6 8.2 12.2




Abundance Ion 163.00 (162.70 to 163.70): E
Ion 194.00 (193.70 to 194.70): E
Ion 164.00 (163.70 to 164.70): E






#51
Acenaphthene
Concen: 21.82 ng
RT: 9.94 min Scan# 1335
Delta R.T. -0.06 min
Lab File: BF099076.D
Acq: 4 Oct 2017 19:36

Tgt Ion:154 Resp: 339952
Ion Ratio Lower Upper
154 100
153 115.4 91.2 136.8
152 54.3 43.8 65.8



Abundance Ion 154.00 (153.70 to 154.70): E
Ion 153.00 (152.70 to 153.70): E
Ion 152.00 (151.70 to 152.70): E









#53
2,4-Dinitrophenol
Concen: 5.93 ng
RT: 9.95 min Scan# 1337
Delta R.T. -0.06 min
Lab File: BF099076.D
Acq: 4 Oct 2017 19:36

Instrument :
BNA_F
ClientSampled :

Tot Ion:184 Resp: 224
Ion Ratio Lower Upper
184 100
63 719.0 54.2 81.2#
154 1110.6 53.5 80.3#




Abundance Ion 184.00 (183.70 to 184.70): E
Ion 63.00 (62.70 to 63.70): BF0
Ion 154.00 (153.70 to 154.70): E






#54
Dibenzofuran
Concen: 15.56 ng
RT: 10.11 min Scan# 1364
Delta R.T. -0.06 min
Lab File: BF099076.D
Acq: 4 Oct 2017 19:36

Tgt Ion:168 Resp: 322841
Ion Ratio Lower Upper
168 100
139 35.9 30.1 45.1
169 12.7 10.9 16.3



Abundance Ion 168.00 (167.70 to 168.70): E
Ion 139.00 (138.70 to 139.70): E
Ion 169.00 (168.70 to 169.70): E









#57
Fluorene
 Concen: 13.35 ng
 RT: 10.45 min Scan# 1422
 Delta R.T. -0.06 min
 Lab File: BF099076.D
 Acq: 4 Oct 2017 19:36

Instrument :
 BNA_F
 ClientSampleId :

Tot Ion:166 Resp: 222647
 Ion Ratio Lower Upper
 166 100
 165 97.8 79.8 119.8
 167 14.4 10.6 16.0




Abundance Ion 166.00 (165.70 to 166.70): E
 Ion 165.00 (164.70 to 165.70): E
 Ion 167.00 (166.70 to 167.70): E






#63
Phenanthrene-d10
 Concen: 20.00 ng
 RT: 11.39 min Scan# 1582
 Delta R.T. -0.05 min
 Lab File: BF099076.D
 Acq: 4 Oct 2017 19:36

Tgt Ion:188 Resp: 424293
 Ion Ratio Lower Upper
 188 100
 94 11.1 8.5 12.7
 80 11.5 9.2 13.8



Abundance Ion 188.00 (187.70 to 188.70): E
 Ion 94.00 (93.70 to 94.70): BF0
 Ion 80.00 (79.70 to 80.70): BF0









#66
4-Bromophenyl-phenylether
Concen: 3.78 ng
RT: 10.70 min Scan# 1464
Delta R.T. -0.29 min
Lab File: BF099076.D
Acq: 4 Oct 2017 19:36

Instrument :
BNA_F
ClientSampled :

Tot Ion:248 Resp: 15705
Ion Ratio Lower Upper
248 100
250 196.1 76.9 115.3#
141 521.2 74.4 111.6#




Abundance Ion 248.00 (247.70 to 248.70): E
Ion 250.00 (249.70 to 250.70): E
Ion 141.00 (140.70 to 141.70): E






#70
Phenanthrene
Concen: 13.82 ng
RT: 11.42 min Scan# 1586
Delta R.T. -0.06 min
Lab File: BF099076.D
Acq: 4 Oct 2017 19:36

Tgt Ion:178 Resp: 313927
Ion Ratio Lower Upper
178 100
176 19.3 15.8 23.8
179 15.4 13.0 19.6



Abundance Ion 178.00 (177.70 to 178.70): E
Ion 176.00 (175.70 to 176.70): E
Ion 179.00 (178.70 to 179.70): E









#71
 Anthracene
 Concen: 2.13 ng
 RT: 11.46 min Scan# 1594
 Delta R.T. -0.06 min
 Lab File: BF099076.D
 Acq: 4 Oct 2017 19:36




Instrument :
 BNA_F
 ClientSampleId :


Tot Ion:178 Resp: 49600
 Ion Ratio Lower Upper
 178 100
 176 20.3 15.4 23.2
 179 16.1 12.6 19.0



#72
 Carbazole
 Concen: 13.75 ng
 RT: 11.62 min Scan# 1621
 Delta R.T. -0.05 min
 Lab File: BF099076.D
 Acq: 4 Oct 2017 19:36

Tgt Ion:167 Resp: 312962
 Ion Ratio Lower Upper
 167 100
 166 22.2 17.6 26.4
 139 13.3 10.9 16.3





#75

Chrysene-d12

Concen: 20.00 ng

RT: 14.03 min Scan# 2030

Delta R.T. -0.06 min

Lab File: BF099076.D

Acq: 4 Oct 2017 19:36

Instrument :

BNA_F

ClientSampleId :


Tot Ion:240 Resp: 255343

Ion Ratio Lower Upper

240 100

120 14.5 9.5 14.3#




236 25.6 20.7 31.1



Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E



#78

Terphenyl-d14

Concen: 93.27 ng

RT: 12.97 min Scan# 1851

Delta R.T. -0.06 min

Lab File: BF099076.D

Acq: 4 Oct 2017 19:36


Tgt Ion:244 Resp: 1047168

Ion Ratio Lower Upper

244 100

212 7.1 5.4 8.0



122 14.8 11.0 16.4






Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E





#86
 Pervlene-d12
 Concen: 20.00 ng
 RT: 15.50 min Scan# 2281
 Delta R.T. -0.07 min
 Lab File: BF099076.D
 Acq: 4 Oct 2017 19:36

Instrument :
 BNA_F
 ClientSampleId :

Tot Ion:	264	Resp:	227270
Ion	Ratio	Lower	Upper
264	100		
260	23.7	19.2	28.8
265	21.5	17.5	26.3

