

Data Path : Z:\HPCHEM1\BNA F\DATA\BF100417\
 Data File : BF099087.D
 Acq On : 5 Oct 2017 00:44
 Operator : SJ/JU
 Sample : I5589-01
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SASP2P-TP-106-T(0-5)

Quant Time: Oct 05 05:45:28 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF092317.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Sep 23 13:50:58 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.86	152	140065	20.00	ng	-0.06
21) Naphthalene-d8	8.15	136	591364	20.00	ng	-0.06
38) Acenaphthene-d10	9.90	164	270127	20.00	ng	-0.06
63) Phenanthrene-d10	11.39	188	470686	20.00	ng	-0.06
75) Chrysene-d12	14.03	240	312900	20.00	ng	-0.06
86) Perylene-d12	15.50	264	237447	20.00	ng	-0.08
System Monitoring Compounds						
5) 2-Fluorophenol	5.49	112	1135805	129.09	ng	-0.05
7) Phenol-d6	6.50	99	1417247	130.57	ng	-0.05
23) Nitrobenzene-d5	7.43	82	851392	92.33	ng	-0.06
41) 2,4,6-Tribromophenol	10.69	330	294515	133.24	ng	-0.06
44) 2-Fluorobiphenyl	9.22	172	1452522	89.77	ng	-0.06
78) Terphenyl-d14	12.97	244	1265378	91.97	ng	-0.06
Target Compounds						
10) Phenol	6.51	94	36760	3.028	ng	93
49) Dimethylphthalate	9.61	163	296593	14.548	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF100417\
Data File : BF099087.D
Acq On : 5 Oct 2017 00:44
Operator : SJ/JU
Sample : I5589-01
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SASP2P-TP-106-T(0-5)

Quant Time: Oct 05 05:45:28 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF092317.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Sep 23 13:50:58 2017
Response via : Initial Calibration

