

Data Path : Z:\HPCHEM1\BNA F\DATA\BF100515\
 Data File : BF082054.D
 Acq On : 5 Oct 2015 19:57
 Operator : UM/IZ
 Sample : G3867-01DL 5X
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SB-07-6.0-7.0-092915DL

Manual Integrations
 APPROVED

MMDadoda
 10/6/2015 2:54:14 PM

Quant Time: Oct 06 05:17:01 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF092815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 05 17:58:12 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.88	152	96990m	20.00	ng	0.00
21) Naphthalene-d8	8.17	136	365903	20.00	ng	0.00
38) Acenaphthene-d10	9.93	164	176728	20.00	ng	0.00
63) Phenanthrene-d10	11.42	188	303694	20.00	ng	0.00
75) Chrysene-d12	14.06	240	277418	20.00	ng	0.00
86) Perylene-d12	15.50	264	239952	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.46	112	161881	30.30	ng	-0.01
7) Phenol-d6	6.50	99	239875	35.68	ng	-0.01
23) Nitrobenzene-d5	7.45	82	124763	22.73	ng	0.00
41) 2,4,6-Tribromophenol	10.72	330	47653	26.62	ng	0.00
44) 2-Fluorobiphenyl	9.25	172	236058	18.77	ng	0.00
78) Terphenyl-d14	13.00	244	228901	20.93	ng	-0.01
Target Compounds						
37) 2-Methylnaphthalene	8.89	142	541526	48.93	ng	# 91
49) Dimethylphthalate	9.65	163	61042	5.00	ng	# 91
57) Fluorene	10.48	166	83104	6.24	ng	# 83
70) Phenanthrene	11.44	178	237079	15.12	ng	# 91
74) Fluoranthene	12.64	202	33372	2.05	ng	# 82
77) Pyrene	12.86	202	42608	2.57	ng	# 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF100515\
Data File : BF082054.D
Acq On : 5 Oct 2015 19:57
Operator : UM/IZ
Sample : G3867-01DL 5X
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SB-07-6.0-7.0-092915DL

Manual Integrations
APPROVED

MMDadoda
10/6/2015 2:54:14 PM

Quant Time: Oct 06 05:17:01 2015
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF092815.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 05 17:58:12 2015
Response via : Initial Calibration

