


Data Path : Z:\HPCHEM1\BNA_F\DATA\BF100516\
 Data File : BF090371.D
 Acq On : 5 Oct 2016 9:19
 Operator : UM/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF100516.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Oct 05 14:57:00 2016



AutoFind: Scans 827, 828, 829; Background Corrected with Scan 824

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	46.9	35036	PASS
68	69	0.00	2	1.6	571	PASS
69	198	0.00	100	48.6	36316	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	57.7	43111	PASS
197	198	0.00	2	0.3	236	PASS
198	198	100	100	100.0	74691	PASS
199	198	5	9	6.8	5088	PASS
275	198	10	60	22.6	16888	PASS
365	198	1	100	2.4	1798	PASS
441	442	0.01	24	16.2	7508	PASS
442	198	50	100	62.2	46456	PASS
443	442	15	24	18.8	8725	PASS