


Data Path : Z:\HPCHEM1\BNA F\DATA\BF100517\
 Data File : BF099128.D
 Acq On : 6 Oct 2017 4:14
 Operator : SJ/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampled :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF092317.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Oct 05 17:20:48 2017



AutoFind: Scans 1622, 1623, 1624; Background Corrected with Scan 1617

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	37.6	20457	PASS
68	69	0.00	2	1.7	312	PASS
69	198	0.00	100	34.3	18654	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	50.9	27677	PASS
197	198	0.00	2	0.7	355	PASS
198	198	100	100	100.0	54389	PASS
199	198	5	9	6.5	3535	PASS
275	198	10	60	22.9	12461	PASS
365	198	1	100	2.5	1366	PASS
441	442	0.01	24	15.9	6763	PASS
442	198	50	100	78.3	42592	PASS
443	442	15	24	19.5	8305	PASS