

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF100721\
 Data File : BF125772.D
 Acq On : 7 Oct 2021 16:53
 Operator : JU/CG
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF100721

Quant Time: Oct 07 17:24:29 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF100721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 07 17:16:23 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	113	0.00
2	1,4-Dioxane	40.000	38.016	5.0	112	0.02
3	Pyridine	40.000	39.180	2.1	111	0.01
4	n-Nitrosodimethylamine	40.000	39.135	2.2	113	0.02
5 S	2-Fluorophenol	80.000	75.476	5.7	111	0.00
6	Aniline	40.000	39.508	1.2	115	0.00
7 S	Phenol-d6	80.000	76.717	4.1	114	0.00
8	2-Chlorophenol	40.000	39.023	2.4	114	0.00
9	Benzaldehyde	40.000	40.178	-0.4	111	0.00
10 C	Phenol	40.000	38.497	3.8	113	0.00
11	bis(2-Chloroethyl)ether	40.000	38.714	3.2	113	0.00
12	1,3-Dichlorobenzene	40.000	38.726	3.2	114	0.00
13 C	1,4-Dichlorobenzene	40.000	38.597	3.5	113	0.00
14	1,2-Dichlorobenzene	40.000	38.818	3.0	114	0.00
15	Benzyl Alcohol	40.000	39.884	0.3	115	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	38.794	3.0	113	0.00
17	2-Methylphenol	40.000	39.022	2.4	115	0.00
18	Hexachloroethane	40.000	39.369	1.6	114	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	39.018	2.5	115	0.00
20	3+4-Methylphenols	40.000	38.521	3.7	113	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	115	0.00
22	Acetophenone	40.000	37.817	5.5	115	0.00
23 S	Nitrobenzene-d5	80.000	78.984	1.3	115	0.00
24	Nitrobenzene	40.000	39.390	1.5	116	0.00
25	Isophorone	40.000	39.313	1.7	116	0.00
26 C	2-Nitrophenol	40.000	43.224	-8.1	119	0.00
27	2,4-Dimethylphenol	40.000	41.327	-3.3	121	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.216	2.0	117	0.00
29 C	2,4-Dichlorophenol	40.000	39.536	1.2	116	0.00
30	1,2,4-Trichlorobenzene	40.000	38.455	3.9	115	0.00
31	Naphthalene	40.000	38.333	4.2	115	0.00
32	Benzoic acid	40.000	45.881	-14.7	125	0.01
33	4-Chloroaniline	40.000	39.505	1.2	117	0.00
34 C	Hexachlorobutadiene	40.000	38.776	3.1	115	0.00
35	Caprolactam	40.000	42.599	-6.5	125	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.695	0.8	117	0.00
37	2-Methylnaphthalene	40.000	38.885	2.8	116	0.00
38	1-Methylnaphthalene	40.000	38.827	2.9	117	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	119	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	37.825	5.4	114	0.00
41 P	Hexachlorocyclopentadiene	40.000	41.096	-2.7	117	0.00
42 S	2,4,6-Tribromophenol	80.000	81.451	-1.8	123	0.00
43 C	2,4,6-Trichlorophenol	40.000	39.383	1.5	119	0.00
44	2,4,5-Trichlorophenol	40.000	39.508	1.2	118	0.00
45 S	2-Fluorobiphenyl	80.000	77.777	2.8	117	0.00
46	1,1'-Biphenyl	40.000	37.879	5.3	116	0.00
47	2-Chloronaphthalene	40.000	38.190	4.5	116	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF100721\
 Data File : BF125772.D
 Acq On : 7 Oct 2021 16:53
 Operator : JU/CG
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF100721

Quant Time: Oct 07 17:24:29 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF100721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 07 17:16:23 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	42.289	-5.7	122	0.00
49	Acenaphthylene	40.000	38.400	4.0	116	0.00
50	Dimethylphthalate	40.000	38.903	2.7	118	0.00
51	2,6-Dinitrotoluene	40.000	41.249	-3.1	121	0.00
52 C	Acenaphthene	40.000	38.706	3.2	118	0.00
53	3-Nitroaniline	40.000	41.663	-4.2	122	0.00
54 P	2,4-Dinitrophenol	40.000	48.710	-21.8	138	0.00
55	Dibenzofuran	40.000	37.962	5.1	116	0.00
56 P	4-Nitrophenol	40.000	43.708	-9.3	126	0.00
57	2,4-Dinitrotoluene	40.000	43.623	-9.1	126	0.00
58	Fluorene	40.000	37.735	5.7	118	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	40.556	-1.4	124	0.00
60	Diethylphthalate	40.000	40.050	-0.1	121	0.00
61	4-Chlorophenyl-phenylether	40.000	37.286	6.8	116	0.00
62	4-Nitroaniline	40.000	43.016	-7.5	127	0.00
63	Azobenzene	40.000	38.892	2.8	118	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	122	0.00
65	4,6-Dinitro-2-methylphenol	40.000	47.996	-20.0	137	0.00
66 c	n-Nitrosodiphenylamine	40.000	37.541	6.1	118	0.00
67	4-Bromophenyl-phenylether	40.000	38.212	4.5	120	0.00
68	Hexachlorobenzene	40.000	38.317	4.2	124	0.00
69	Atrazine	40.000	40.696	-1.7	127	0.00
70 C	Pentachlorophenol	40.000	43.732	-9.3	130	0.00
71	Phenanthrene	40.000	37.394	6.5	119	0.00
72	Anthracene	40.000	38.215	4.5	122	0.00
73	Carbazole	40.000	39.252	1.9	125	0.00
74	Di-n-butylphthalate	40.000	40.861	-2.2	128	0.00
75 C	Fluoranthene	40.000	40.778	-1.9	132	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	155	0.00
77	Benzydine	40.000	36.913	7.7	142	0.00
78	Pyrene	40.000	35.422	11.4	135	0.00
79 S	Terphenyl-d14	80.000	71.882	10.1	136	0.00
80	Butylbenzylphthalate	40.000	46.410	-16.0	179	0.00
81	Benzo(a)anthracene	40.000	37.857	5.4	150	0.00
82	3,3'-Dichlorobenzidine	40.000	38.737	3.2	144	0.00
83	Chrysene	40.000	38.448	3.9	152	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	48.724	-21.8	186	0.00
85 c	Di-n-octyl phthalate	40.000	38.603	3.5	137	0.00
86 I	Perylene-d12	20.000	20.000	0.0	104	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	38.970	2.6	102	0.00
88	Benzo(b)fluoranthene	40.000	41.995	-5.0	110	0.00
89	Benzo(k)fluoranthene	40.000	41.360	-3.4	110	0.00
90 C	Benzo(a)pyrene	40.000	40.877	-2.2	106	0.00
91	Dibenzo(a,h)anthracene	40.000	39.096	2.3	101	0.00
92	Benzo(g,h,i)perylene	40.000	38.664	3.3	101	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF100721\
Data File : BF125772.D
Acq On : 7 Oct 2021 16:53
Operator : JU/CG
Sample : SSTDICV040
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF100721

Quant Time: Oct 07 17:24:29 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF100721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 07 17:16:23 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0