

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF100722\
 Data File : BF130594.D
 Acq On : 08 Oct 2022 01:43
 Operator : CG\JU
 Sample : N5020-01DL 50X
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 TR-5-100522DL

Quant Time: Oct 10 01:48:15 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 07 02:06:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.616	152	68783	20.000	ng	0.00
21) Naphthalene-d8	7.892	136	234860	20.000	ng	-0.01
39) Acenaphthene-d10	9.645	164	117414	20.000	ng	0.00
64) Phenanthrene-d10	11.122	188	218697	20.000	ng	-0.01
76) Chrysene-d12	13.757	240	200683	20.000	ng	0.00
86) Perylene-d12	15.139	264	136767	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.222	112	9973	2.515	ng	0.00
7) Phenol-d6	6.269	99	12151	2.449	ng	0.00
23) Nitrobenzene-d5	7.181	82	7090	1.542	ng	-0.01
42) 2,4,6-Tribromophenol	10.433	330	2633	2.340	ng	0.00
45) 2-Fluorobiphenyl	8.969	172	14026	1.740	ng	-0.01
79) Terphenyl-d14	12.704	244	17344	1.432	ng	-0.01
Target Compounds						
80) Butylbenzylphthalate	13.192	149	348633	53.605	ng	Qvalue 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF100722\
Data File : BF130594.D
Acq On : 08 Oct 2022 01:43
Operator : CG\JU
Sample : N5020-01DL 50X
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
TR-5-100522DL

Quant Time: Oct 10 01:48:15 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091422.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 07 02:06:46 2022
Response via : Initial Calibration

