


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF100722\
 Data File : BF130567.D
 Acq On : 07 Oct 2022 09:34
 Operator : CG\JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091422.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Oct 10 13:41:25 2022






AutoFind: Scans 1582, 1583, 1584; Background Corrected with Scan 1577

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	34.5	41520	PASS
68	69	0.00	2	1.8	788	PASS
69	198	0.00	100	36.6	43975	PASS
70	69	0.00	2	0.5	220	PASS
127	198	10	80	50.7	61029	PASS
197	198	0.00	2	0.6	667	PASS
198	198	100	100	100.0	120267	PASS
199	198	5	9	6.5	7803	PASS
275	198	10	60	29.6	35587	PASS
365	198	1	100	4.3	5196	PASS
441	198	0.01	100	14.9	17945	PASS
442	442	50	100	100.0	119437	PASS
443	442	15	24	18.3	21853	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF100722\
 Data File : BF130567.D
 Acq On : 07 Oct 2022 09:34
 Operator : CG\JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Quant Time: Oct 10 04:10:02 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 07 02:06:46 2022
 Response via : Initial Calibration



TIC: BF130567.D\data.ms

(70) Pentachlorophenol (C)

10.945min (-0.000) 24753.44 ng




response 98182

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	63.10	64.74
264.00	63.10	61.82
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF100722\
Data File : BF130567.D
Acq On : 07 Oct 2022 09:34
Operator : CG\JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Oct 10 04:10:02 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091422.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 07 02:06:46 2022
Response via : Initial Calibration



TIC: BF130567.D\data.ms

(77) Benzidine

12.474min (+ 0.006) 0.00 ng

response 379038

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.70	13.64
183.00	11.40	11.48
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
10/7/2022	BNA_F	<u>BF130567.D</u>
Compound Name	Response	Retention Time
DDT	209291	13.292
DDD	5924	12.992
DDE	808	12.657
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
6732	216023	3.12

Instrument :
BNA_F
ClientSampleId :
DFTPP