



Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF100724\
Data File : BF139676.D
Acq On : 07 Oct 2024 09:50
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF100124.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Wed Oct 02 04:58:47 2024






AutoFind: Scans 1622, 1623, 1624; Background Corrected with Scan 1616

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	48.6	53317	PASS
68	69	0.00	2	1.8	799	PASS
69	198	0.00	100	40.7	44619	PASS
70	69	0.00	2	0.6	254	PASS
127	198	10	80	48.6	53251	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	109611	PASS
199	198	5	9	6.5	7165	PASS
275	198	10	60	26.3	28832	PASS
365	198	1	100	3.4	3704	PASS
441	198	0.01	100	13.3	14554	PASS
442	442	50	100	100.0	91557	PASS
443	442	15	24	19.3	17668	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF100724\
Data File : BF139676.D
Acq On : 07 Oct 2024 09:50
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Oct 07 11:58:16 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF100124.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 02 04:58:47 2024
Response via : Initial Calibration



TIC: BF139676.D\data.ms

(70) Pentachlorophenol (C)

11.198min (0.000) 30101.84 ng




response 128882

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.50	62.10
264.00	61.60	62.66
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF100724\
Data File : BF139676.D
Acq On : 07 Oct 2024 09:50
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Oct 07 11:58:16 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF100124.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 02 04:58:47 2024
Response via : Initial Calibration



TIC: BF139676.D\data.ms

(77) Benzidine

12.727min (+ 0.000) 0.00 ng

response 640411

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.20	14.10
183.00	11.00	11.20
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
10/7/2024	BNA_F	<u>BF139676.D</u>
Compound Name	Response	Retention Time
DDT	351115	13.551
DDD	6258	13.28
DDE	233	12.91
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
6491	357606	1.82

Instrument :
BNA_F
ClientSampleId :
DFTPP