

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF100824\
 Data File : BF139701.D
 Acq On : 08 Oct 2024 09:39
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 08 11:37:13 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF100124.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 02 04:58:47 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	82	0.00
2	1,4-Dioxane	40.000	38.731	3.2	81	-0.02
3	Pyridine	40.000	40.649	-1.6	87	-0.01
4	n-Nitrosodimethylamine	40.000	38.472	3.8	80	-0.02
5 S	2-Fluorophenol	80.000	80.332	-0.4	85	0.00
6	Aniline	40.000	36.017	10.0	79	0.00
7 S	Phenol-d6	80.000	80.236	-0.3	85	0.00
8	2-Chlorophenol	40.000	40.056	-0.1	85	0.00
9	Benzaldehyde	40.000	38.007	5.0	81	0.00
10 C	Phenol	40.000	40.414	-1.0	84	0.00
11	bis(2-Chloroethyl)ether	40.000	36.853	7.9	80	0.00
12	1,3-Dichlorobenzene	40.000	39.402	1.5	83	0.00
13 C	1,4-Dichlorobenzene	40.000	39.134	2.2	84	0.00
14	1,2-Dichlorobenzene	40.000	39.055	2.4	83	0.00
15	Benzyl Alcohol	40.000	39.200	2.0	81	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	37.501	6.2	79	0.00
17	2-Methylphenol	40.000	39.395	1.5	82	0.00
18	Hexachloroethane	40.000	38.304	4.2	81	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	37.539	6.2	80	0.00
20	3+4-Methylphenols	40.000	39.616	1.0	84	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	82	0.00
22	Acetophenone	40.000	38.365	4.1	83	0.00
23 S	Nitrobenzene-d5	80.000	77.392	3.3	82	0.00
24	Nitrobenzene	40.000	38.719	3.2	82	0.00
25	Isophorone	40.000	37.551	6.1	80	0.00
26 C	2-Nitrophenol	40.000	39.695	0.8	80	0.00
27	2,4-Dimethylphenol	40.000	38.645	3.4	83	0.00
28	bis(2-Chloroethoxy)methane	40.000	37.544	6.1	80	0.00
29 C	2,4-Dichlorophenol	40.000	39.547	1.1	83	0.00
30	1,2,4-Trichlorobenzene	40.000	38.544	3.6	82	0.00
31	Naphthalene	40.000	38.983	2.5	83	0.00
32	Benzoic acid	40.000	42.015	-5.0	82	0.00
33	4-Chloroaniline	40.000	37.175	7.1	77	0.00
34 C	Hexachlorobutadiene	40.000	37.826	5.4	81	0.00
35	Caprolactam	40.000	40.638	-1.6	87	0.00
36 C	4-Chloro-3-methylphenol	40.000	38.953	2.6	84	0.00
37	2-Methylnaphthalene	40.000	37.859	5.4	81	0.00
38	1-Methylnaphthalene	40.000	38.306	4.2	82	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	83	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	37.938	5.2	82	0.00
41 P	Hexachlorocyclopentadiene	40.000	33.111	17.2	64	0.00
42 S	2,4,6-Tribromophenol	80.000	75.802	5.2	81	0.00
43 C	2,4,6-Trichlorophenol	40.000	37.361	6.6	79	0.00
44	2,4,5-Trichlorophenol	40.000	38.548	3.6	83	0.00
45 S	2-Fluorobiphenyl	80.000	74.382	7.0	82	0.00
46	1,1'-Biphenyl	40.000	37.739	5.7	82	0.00
47	2-Chloronaphthalene	40.000	38.053	4.9	82	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF100824\
 Data File : BF139701.D
 Acq On : 08 Oct 2024 09:39
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 08 11:37:13 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF100124.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 02 04:58:47 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	39.535	1.2	84	0.00
49	Acenaphthylene	40.000	38.268	4.3	83	0.00
50	Dimethylphthalate	40.000	37.804	5.5	82	0.00
51	2,6-Dinitrotoluene	40.000	39.597	1.0	84	0.00
52 C	Acenaphthene	40.000	38.002	5.0	82	0.00
53	3-Nitroaniline	40.000	39.224	1.9	84	0.00
54 P	2,4-Dinitrophenol	40.000	40.050	-0.1	81	0.00
55	Dibenzofuran	40.000	38.138	4.7	83	0.00
56 P	4-Nitrophenol	40.000	37.603	6.0	79	0.01
57	2,4-Dinitrotoluene	40.000	39.799	0.5	84	0.00
58	Fluorene	40.000	38.592	3.5	85	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	38.860	2.9	84	0.00
60	Diethylphthalate	40.000	37.701	5.7	82	0.00
61	4-Chlorophenyl-phenylether	40.000	37.627	5.9	84	0.00
62	4-Nitroaniline	40.000	39.364	1.6	85	0.00
63	Azobenzene	40.000	37.325	6.7	80	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	86	0.00
65	4,6-Dinitro-2-methylphenol	40.000	40.835	-2.1	83	0.00
66 c	n-Nitrosodiphenylamine	40.000	37.659	5.9	84	0.00
67	4-Bromophenyl-phenylether	40.000	36.957	7.6	82	0.00
68	Hexachlorobenzene	40.000	37.307	6.7	83	0.00
69	Atrazine	40.000	28.750	28.1#	74	0.00
70 C	Pentachlorophenol	40.000	37.485	6.3	78	0.00
71	Phenanthrene	40.000	38.107	4.7	85	0.00
72	Anthracene	40.000	38.465	3.8	86	0.00
73	Carbazole	40.000	39.740	0.6	88	0.00
74	Di-n-butylphthalate	40.000	38.472	3.8	84	0.00
75 C	Fluoranthene	40.000	39.434	1.4	88	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	90	0.00
77	Benzydine	40.000	31.881	20.3	73	0.00
78	Pyrene	40.000	39.700	0.7	88	0.00
79 S	Terphenyl-d14	80.000	78.134	2.3	89	0.00
80	Butylbenzylphthalate	40.000	41.357	-3.4	92	0.00
81	Benzo(a)anthracene	40.000	38.481	3.8	91	0.00
82	3,3'-Dichlorobenzidine	40.000	36.202	9.5	83	0.00
83	Chrysene	40.000	39.170	2.1	89	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	41.094	-2.7	91	0.00
85 c	Di-n-octyl phthalate	40.000	37.624	5.9	84	0.00
86 I	Perylene-d12	20.000	20.000	0.0	76	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	40.421	-1.1	75	0.01
88	Benzo(b)fluoranthene	40.000	36.977	7.6	79	0.00
89	Benzo(k)fluoranthene	40.000	41.715	-4.3	77	0.00
90 C	Benzo(a)pyrene	40.000	39.168	2.1	77	0.00
91	Dibenzo(a,h)anthracene	40.000	40.333	-0.8	74	0.01
92	Benzo(g,h,i)perylene	40.000	41.820	-4.6	77	0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF100824\
Data File : BF139701.D
Acq On : 08 Oct 2024 09:39
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Oct 08 11:37:13 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF100124.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 02 04:58:47 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0