

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF100825\
 Data File : BF143899.D
 Acq On : 08 Oct 2025 17:07
 Operator : RC/JU
 Sample : Q3266-02
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ENV-4-6FT

Quant Time: Oct 08 17:28:18 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF100625.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 06 15:29:47 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.881	152	95054	20.000	ng	0.00
21) Naphthalene-d8	8.163	136	368458	20.000	ng	0.00
39) Acenaphthene-d10	9.922	164	204631	20.000	ng	0.00
64) Phenanthrene-d10	11.410	188	351450	20.000	ng	0.00
76) Chrysene-d12	14.057	240	205903	20.000	ng	-0.01
86) Perylene-d12	15.533	264	263892	20.000	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.493	112	498414	83.266	ng	0.00
7) Phenol-d6	6.498	99	602460	84.439	ng	-0.02
23) Nitrobenzene-d5	7.440	82	366542	60.439	ng	-0.02
42) 2,4,6-Tribromophenol	10.710	330	160278	78.638	ng	-0.01
45) 2-Fluorobiphenyl	9.239	172	727899	56.442	ng	-0.01
79) Terphenyl-d14	12.998	244	684348	56.804	ng	-0.01


Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF100825\
Data File : BF143899.D
Acq On : 08 Oct 2025 17:07
Operator : RC/JU
Sample : Q3266-02
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ENV-4-6FT

Quant Time: Oct 08 17:28:18 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF100625.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 06 15:29:47 2025
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.881 min Scan# 797
Delta R.T. -0.005 min
Lab File: BF143899.D
Acq: 08 Oct 2025 17:07




Instrument :
BNA_F
ClientSampleId :
ENV-4-6FT

Tgt Ion:152 Resp: 95054
Ion Ratio Lower Upper
152 100
150 156.6 124.4 186.6
115 63.0 48.1 72.1



#5
2-Fluorophenol
Concen: 83.266 ng
RT: 5.493 min Scan# 561
Delta R.T. -0.006 min
Lab File: BF143899.D
Acq: 08 Oct 2025 17:07

Tgt Ion:112 Resp: 498414
Ion Ratio Lower Upper
112 100
64 62.1 49.1 73.7
63 29.0 24.5 36.7









#7
Phenol-d6
Concen: 84.439 ng
RT: 6.498 min Scan# 731
Delta R.T. -0.017 min
Lab File: BF143899.D
Acq: 08 Oct 2025 17:07




Instrument :
BNA_F
ClientSampleId :
ENV-4-6FT

Tgt Ion: 99 Resp: 602460
Ion Ratio Lower Upper
99 100
42 25.2 21.0 31.4
71 33.0 27.8 41.8



#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.163 min Scan# 1015
Delta R.T. -0.006 min
Lab File: BF143899.D
Acq: 08 Oct 2025 17:07

Tgt Ion: 136 Resp: 368458
Ion Ratio Lower Upper
136 100
137 10.7 8.5 12.7
54 10.7 8.2 12.4
68 7.6 6.2 9.4









#23
Nitrobenzene-d5
Concen: 60.439 ng
RT: 7.440 min Scan# 893
Delta R.T. -0.017 min
Lab File: BF143899.D
Acq: 08 Oct 2025 17:07




Instrument :
BNA_F
ClientSampleId :
ENV-4-6FT


Tgt Ion: 82 Resp: 366542
Ion Ratio Lower Upper
82 100
128 41.5 34.0 51.0
54 63.9 53.2 79.8



#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.922 min Scan# 1314
Delta R.T. -0.005 min
Lab File: BF143899.D
Acq: 08 Oct 2025 17:07

Tgt Ion:164 Resp: 204631
Ion Ratio Lower Upper
164 100
162 98.6 80.2 120.2
160 44.1 35.4 53.0









#42
2,4,6-Tribromophenol
Concen: 78.638 ng
RT: 10.710 min Scan# 1448
Delta R.T. -0.012 min
Lab File: BF143899.D
Acq: 08 Oct 2025 17:07




Instrument :
BNA_F
ClientSampleId :
ENV-4-6FT


Tgt Ion:330 Resp: 160278
Ion Ratio Lower Upper
330 100
332 97.3 76.9 115.3
141 26.1 20.2 30.4



#45
2-Fluorobiphenyl
Concen: 56.442 ng
RT: 9.239 min Scan# 1198
Delta R.T. -0.012 min
Lab File: BF143899.D
Acq: 08 Oct 2025 17:07

Tgt Ion:172 Resp: 727899
Ion Ratio Lower Upper
172 100
171 34.2 27.4 41.0
170 23.0 18.6 28.0









#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.410 min Scan# 11
Delta R.T. -0.006 min
Lab File: BF143899.D
Acq: 08 Oct 2025 17:07




Instrument :
BNA_F
ClientSampleId :
ENV-4-6FT


Tgt Ion:188 Resp: 351450
Ion Ratio Lower Upper
188 100
94 10.1 7.8 11.6
80 10.0 7.9 11.9



#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.057 min Scan# 2017
Delta R.T. -0.011 min
Lab File: BF143899.D
Acq: 08 Oct 2025 17:07

Tgt Ion:240 Resp: 205903
Ion Ratio Lower Upper
240 100
120 10.9 8.6 13.0
236 25.9 21.1 31.7









#79
 Terphenyl-d14
 Concen: 56.804 ng
 RT: 12.998 min Scan# 1837
 Delta R.T. -0.012 min
 Lab File: BF143899.D
 Acq: 08 Oct 2025 17:07

Instrument :
 BNA_F
 ClientSampleId :
 ENV-4-6FT

Tgt Ion:244 Resp: 684348
 Ion Ratio Lower Upper
 244 100
 212 6.4 5.4 8.0
 122 9.2 6.4 9.6



#86
 Perylene-d12
 Concen: 20.000 ng
 RT: 15.533 min Scan# 2268
 Delta R.T. -0.012 min
 Lab File: BF143899.D
 Acq: 08 Oct 2025 17:07

Tgt Ion:264 Resp: 263892
 Ion Ratio Lower Upper
 264 100
 260 21.2 17.8 26.8
 265 21.6 17.5 26.3

