


Data Path : Z:\HPCHEM1\BNA F\Data\BF100917\
 Data File : BF099285.D
 Acq On : 10 Oct 2017 2:25
 Operator : SJ/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF092317.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Oct 05 17:20:48 2017



AutoFind: Scans 1621, 1622, 1623; Background Corrected with Scan 1617

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	34.6	21047	PASS
68	69	0.00	2	1.0	208	PASS
69	198	0.00	100	33.7	20476	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	49.0	29812	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	60800	PASS
199	198	5	9	6.4	3890	PASS
275	198	10	60	22.4	13641	PASS
365	198	1	100	2.1	1251	PASS
441	442	0.01	24	15.4	7518	PASS
442	198	50	100	80.2	48787	PASS
443	442	15	24	18.7	9137	PASS