

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF101015\
 Data File : BF082209.D
 Acq On : 11 Oct 2015 7:39
 Operator : UM/IZ
 Sample : G3889-02DL 10X
 Misc :
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 MGP123-5-7DL

Manual Integrations
 APPROVED

MMDadoda
 10/12/2015 2:38:55 PM

Quant Time: Oct 12 04:23:02 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF101015.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 12 02:59:50 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.83	152	92597	20.00	ng	0.00
21) Naphthalene-d8	8.13	136	368127	20.00	ng	0.00
38) Acenaphthene-d10	9.89	164	166545	20.00	ng	0.00
63) Phenanthrene-d10	11.37	188	283069	20.00	ng	0.00
75) Chrysene-d12	14.00	240	249873	20.00	ng	-0.01
86) Perylene-d12	15.44	264	234807	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.43	112	53089	11.57	ng	0.00
7) Phenol-d6	6.47	99	75866	12.71	ng	-0.01
23) Nitrobenzene-d5	7.41	82	44356	8.09	ng	-0.01
41) 2,4,6-Tribromophenol	10.67	330	10328	6.61	ng	0.00
44) 2-Fluorobiphenyl	9.20	172	112206	11.17	ng	-0.01
78) Terphenyl-d14	12.95	244	92315	9.79	ng	-0.01
Target Compounds						Qvalue
31) Naphthalene	8.15	128	766227	48.02	ng	99
37) 2-Methylnaphthalene	8.83	142	126428	11.43	ng	98
49) Dimethylphthalate	9.60	163	27183	2.32	ng	# 99
51) Acenaphthene	9.91	154	39503	4.23	ng	99
57) Fluorene	10.42	166	26962m	2.56	ng	
70) Phenanthrene	11.39	178	92534	6.67	ng	97
71) Anthracene	11.44	178	35749	2.50	ng	97
77) Pyrene	12.81	202	40945	2.76	ng	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF101015\
Data File : BF082209.D
Acq On : 11 Oct 2015 7:39
Operator : UM/IZ
Sample : G3889-02DL 10X
Misc :
ALS Vial : 42 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
MGP123-5-7DL

Manual Integrations
APPROVED

MMDadoda
10/12/2015 2:38:55 PM

Quant Time: Oct 12 04:23:02 2015
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF101015.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 12 02:59:50 2015
Response via : Initial Calibration

