





Method Name: Z:\HPCHEM1\BNA_F\METHODS\8270-BF101015.M
Calibration Table Last Updated: Mon Oct 12 01:46:21 2015

Hexachlorocyclopentadiene

Response Ratio




Resp Ratio = $3.36 \times 10^{-1} \times \text{Amt} - 1.39 \times 10^{-1}$
Coef of Det (r^2) = 0.994 Curve Fit: Linear

Method Name: Z:\HPCHEM1\BNA_F\METHODS\8270-BF101015.M
Calibration Table Last Updated: Mon Oct 12 01:46:21 2015

2,4-Dinitrophenol

Response Ratio



Resp Ratio = $1.80 \times 10^{-1} \times \text{Amt} - 5.77 \times 10^{-2}$
Coef of Det (r^2) = 0.996 Curve Fit: Linear

Method Name: Z:\HPCHEM1\BNA_F\METHODS\8270-BF101015.M
Calibration Table Last Updated: Mon Oct 12 01:46:21 2015